

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: MAAYON SUPER HEALTH CENTER ADDRESS: MAAYON, CAPIZ
SERVICE REQUESTED BY: END USER DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: FXRD-3643VAW SERIAL NO.: VEDBEB102

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS UNDER WARRANTY

COMPLAINT: PREVENTIVE MAINTENANCE SERVICE WARRANTY (2ND VISIT OF 1ST YEAR)/LINES ON SCREEN

ACTIONS TAKEN

-CHECKED MONITOR FOR ANY SIGNS OF PHYSICAL DAMAGE, SUCH AS ACCIDENTAL DROPPING; NON VISIBLE DEFECTS
-TESTED USING A DIFFERENT CPU TO ISOLATE THE ISSUE; LINES ON THE DISPLAY REMAINED, CONFIRMING THAT THE PROBLEM ORIGINATES FROM THE MONITOR ITSELF/END-USER PROVIDE EXTRA MONITOR
-THE DELL SERVICE CENTER WAS CONTACTED TO INITIATE A WARRANTY CLAIM FOR THE DEFECTIVE UNIT
-PERFORMED THE 1ST YEAR-2ND VISIT UNDER THE 2-YEAR SEMI ANNUAL PREVENTIVE MAINTENANCE WARRANTY.
-PERFORMED GENERAL CLEANING
-CHECKED FACILITY OUTLET POWER SUPPLY; 227.2VAC
-CHECKED UPS POWER OUTPUT; 226.5VAC
-CHECKED PC'S C-MOS BATTERY; 3.061VDC
-CHECKED FPD CRADLE CHARGER SLOT A AND SLOT B; ALL WORKING
-CHECKED BATTERY FPD CHARGING STATUS; OK
-CHECKED FPD TOTAL SHOT COUNT; 135 SHOTS
-CHECKED FPD SHOCKLOG; NO HISTORY DETECTED
-CREATED SYSTEM RESTORATION: SUCCESS
-PERFORMED FPD GAIN CALIBRATION; DONE
-PERFORMED IMAGE ACQUISITION; OK
-TESTED THE UNIT

REMARKS/RECOMMENDATIONS

CURRENTLY AWAITING THE DELIVERY OF THE REPLACEMENT MONITOR FROM DELL/UNIT IS GOOD WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED


ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 6, 2026	12:30 P.M	4:50 P.M	SALES INVOICE
			PURCHASE ORDER
			UNDER WARRANTY

SERVICED BY:


JOMARIE T. RANAFLOR
NAME & SIGNATURE

ACKNOWLEDGED BY:


DR. SANDRA SORRONGON
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

maayonrhu@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting