

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DAVAO DOCTORS HOSPITAL - DUMOY ADDRESS: DAVAO CITY
SERVICE REQUESTED BY: _____ DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: UD150L-30FX SERIAL NO.: 0162R36901

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS _____

COMPLAINT:**ACTIONS TAKEN**

SCHEDULED PREVENTIVE MAINTENANCE UNDER SERVICE CONTRACT - 2ND VISIT

- Checked and cleaned the unit
- Checked other mechanical movements
- Checked collimator functions, adjusted collimator knob (tightened)
- Checked DR System, check system unit and DR software
- Checked and cleaned boards and fuses
- Checked and cleaned stand and table
- Performed X-ray tube calibration
- Tested image acquisition and exposure

- Endorsed to end-user and biomed

REMARKS/RECOMMENDATIONS

UNIT IN WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/06/2026	11:00PM	12:30PM	SALES INVOICE _____
			PURCHASE ORDER _____ AS PER S.C.

SERVED BY:


DANIEL N. P. / KENT MONTEVERDE
NAME & SIGNATURE

ACKNOWLEDGED BY:


JONIEL ANSEN TESARA
NAME & SIGNATURE

CONTACT NO: _____

EMAIL ADD: _____

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting