

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ST. ANTHONY MOTHER AND CHILD HOSPITAL ADDRESS: CEBU CITY, CEBU
SERVICE REQUESTED BY: SPC DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED MF WITH VIEWWORKS SERIAL NO.: V5DBJV062

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☒ OTHERS

COMPLAINT: PACS/RIS CONCERN

ACTIONS TAKEN

-END USER REPORTED ANNOTATION AND COLLIMATION NOT APPEARING DURING SENDING TO PACS SERVER
-ENABLED ANNOTATION AND COLLIMATION WHEN SENDING TO PACS
-TESTED SENDING IMAGES TO BOTH VOYAGER AND LIFETRACK PACS; SUCCESS
-INTEGRATED MWL/RIS TO VIEWWORKS
-TESTED CONNECTION; SUCCESS
-TESTED QUERY OF MWL; SUCCESS
-TESTED IMAGE OUTPUT; NORMAL
-ENDORSED TO END USER
-ENDORSED ALSO PMS CONTRACT TO END USER; RECOMMEND TO AVAIL PMS CONTRACT

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION AND READY TO USE

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 05, 2026	11:00 AM	2:00 PM	SALES INVOICE
			PURCHASE ORDER
			FREE OF CHARGE

SERVED BY:


JASPER B. REDOBLE
NAME & SIGNATURE

ACKNOWLEDGED BY:


KAREN TAMPUS, RRT
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting