

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: TAPAZ SUPER HEALTH CENTER ADDRESS: TAPAZ, CAPIZ
SERVICE REQUESTED BY: END USER DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: FXRD-3643VAW SERIAL NO.: VEDBEB113

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: PREVENTIVE MAINTENANCE SERVICE WARRANTY (2ND VISIT OF 1ST YEAR)

ACTIONS TAKEN

-PERFORMED THE 1ST YEAR-2ND VISIT UNDER THE 2-YEAR SEMI ANNUAL PREVENTIVE MAINTENANCE WARRANTY.
-PERFORMED GENERAL CLEANING
-CHECKED FACILITY OUTLET POWER SUPPLY; 247.9VAC
-CHECKED UPS POWER OUTPUT; 209.6VAC
-CHECKED PC'S C-MOS BATTERY; 3.039VDC
-CHECKED BATTERY FPD CHARGING STATUS; OK
-CHECKED FPD TOTAL SHOT COUNT; 52 SHOTS
-CHECKED FPD SHOCKLOG; NO HISTORY DETECTED
-CREATED SYSTEM RESTORATION: SUCCESS
-PERFORMED FPD GAIN CALIBRATION; DONE
-PERFORMED IMAGE AQUISITION; OK
-TESTED THE UNIT

REMARKS/RECOMMENDATIONS

UNIT IS GOOD WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS


☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED


ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 5, 2026	9:30 A.M	12:20 P.M	SALES INVOICE
			PURCHASE ORDER
			UNDER WARRANTY

SERVED BY:


JOMARIE T. RENAFLOR
NAME & SIGNATURE

ACKNOWLEDGED BY:


APRILYN V. LARROZA
NAME & SIGNATURE

CONTACT NO:**EMAIL ADD:**

larrozaaprillyn@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting