

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DAVAO DOCTORS HOSPITAL ADDRESS: DAVAO CITY
SERVICE REQUESTED BY: DEPARTMENT: XRAY ROOM #2
EQUIPMENT/MODEL: MX8 SERIAL NO.: MQ00005A5042

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: SCHEDULED PREVENTIVE MAINTENANCE UNDER SERVICE CONTRACT

ACTIONS TAKEN**PERFORMED PREVENTIVE MAINTENANCE**

-checked and cleaned whole unit
-checked handle for tube rotation, loosen
-removed covers, retightened screws of handle and applied lock tite, tested : good
-checked and cleaned boards and fuses
-checked 9volts, CMOS and 20pcs batteries: good
-checked batteries detection both (+) and (-) side: good
-checked voltage drop check mode: good
-checked mechanical movements, applied WD40
-checked wire rope
-checked MWL and DICOM sending: good
-checked DR system, checked FPD: good
-performed xray tube calibration: good
-tested xray and image acquisition: good
-exposure count= 19,761, already reset

REMARKS/RECOMMENDATIONS

UNIT OPERATIONAL

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 05, 2026	9:00AM	11:00AM	SALES INVOICE
			PURCHASE ORDER
			SC

SERVED BY:


DANIEL NOPPE
NAME & SIGNATURE

ACKNOWLEDGED BY:


JONIEL JANSEN TESARA
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting