

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: CEBU VETERINARY DOCTORS ADDRESS: CEBU CITY, CEBU
SERVICE REQUESTED BY: SPC DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: MUX-10 SERIAL NO.: 3YCFC393B010

SERVICE CATEGORY

☐ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: F12 (KV UNDER)

ACTIONS TAKEN

-CONDUCTED VISIT
-PERFORMED FURTHER TROUBLESHOOTING FOR F12 ERROR
-CHECKED KVFB- AND KVFB+ WAVEFORMS @ 80 kVp 10 mAs
-NO FEEDBACK WAVEFORM
-DOUBLE CHECKED CABLE CONNECTION SET UP
-PERFORMED EXPOSURE @80 kVp 5 mAs TO CHECK FEEDBACK WAVEFORM; NO FEEDBACK WAVEFORM
-PERFORMED EXPOSURE @90 kVp 5 mAs; F12 ERROR APPEARING
-WHEN PERFORMING EXPOSURE @80 kVp 5 mAs; THE FEEDBACK KVP AT DOSEMETER IS ONLY 39.60~ kVp
-ENDORSED TO CLIENT

REMARKS/RECOMMENDATIONS

UNIT IS WORKING BUT WITH LIMITED EXPOSURE FUNCTIONS DUE TO F12 ERROR.

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☒ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 08, 2026	11:00 AM	1:00 PM	SALES INVOICE
			PURCHASE ORDER

SERVED BY:


JASPER B. REDOBLE
NAME & SIGNATURE

ACKNOWLEDGED BY:


FELUCHE GIMENA
NAME & SIGNATURE

CONTACT NO:
EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting