

- ☐ Analytical  
☒ Medical  
☐ Testing Machine/Big Range  
☐ Balance

**TECHNICAL SERVICE REPORT**

CUSTOMER NAME: MADONNA MEDICAL CENTER, INC.

ADDRESS: KIDAPAWAN CITY

SERVICE REQUESTED BY:

DEPARTMENT: RADIOLOGY

EQUIPMENT/MODEL: MobileArt eco MUX-10 with DR

SERIAL NO.: 3YCF7FE9001

## SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

## COMPLAINT:

## ACTIONS TAKEN

## SCHEDULED PREVENTIVE MAINTENANCE UNDER WARRANTY - LAST VISIT

- Checked and cleaned unit
- Checked Wire rope, OK
- Checked mechanical movements, applied lubricant to moving parts
- Checked fuses and CMOS battery, 3.063VDC
- Applied contact cleaner to board wire connections
- Performed x-ray tube aging/warm-up
- Calibrated unit's FVR (Filament Voltage Radiography)
- Verified overall tube exposure count: 553
- Checked and cleaned DR System: both Acquisition and Reading PC
- Checked UPS: OK
- Checked and cleaned SCU & FPD: OK
- Tested exposure & image acquisition, OK
- Tested image sending, OK

- Endorsed to end-user

## REMARKS/RECOMMENDATIONS

UNIT IS IN GOOD WORKING CONDITION

## SUBMITTED ORIGINAL DOCUMENTS

- |                                         |                                                     |                                          |                                            |
|-----------------------------------------|-----------------------------------------------------|------------------------------------------|--------------------------------------------|
| <input type="checkbox"/> STICKER        | <input type="checkbox"/> CALIBRATION CERTIFICATE    | <input type="checkbox"/> PM CERTIFICATE  | <input type="checkbox"/> COMPLETION REPORT |
| <input type="checkbox"/> SYSTEM CHECK   | <input type="checkbox"/> IQ DOCUMENTS               | <input type="checkbox"/> OQ DOCUMENTS    | <input type="checkbox"/> PM DOCUMENTS      |
| <input type="checkbox"/> SERVICE REPORT | <input type="checkbox"/> CALIBRATION DATA (MEDICAL) | <input type="checkbox"/> INVOICE/BILLING | <input type="checkbox"/> DELIVERY RECEIPT  |


## PARTS SUPPLIED/RECOMMENDED

| ITEM NO. | PART NUMBER | DESCRIPTION | QTY | DELIVERY RECEIPT |
|----------|-------------|-------------|-----|------------------|
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |

| DATE OF SERVICE | TIME STARTED | TIME FINISHED | BILLING STATEMENT |
|-----------------|--------------|---------------|-------------------|
| 05/08/2026      | 9:00AM       | 1:00PM        | SALES INVOICE     |
|                 |              |               | PURCHASE ORDER    |
|                 |              |               | AS PER W.         |

SERVICED BY:

ACKNOWLEDGED BY:

  
KENT P. MONTEVERDE  
NAME & SIGNATURE  
GERRIEN GRACE G. VILBAR, RRT  
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

mmciradiologydept2020@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting