

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DAVAO DOCTORS HOSPITAL ADDRESS: DAVAO CITY
SERVICE REQUESTED BY: DEPARTMENT: XRAY ROOM #2
EQUIPMENT/MODEL: MX7 SERIAL NO.: MPF168167004

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: SCHEDULED PREVENTIVE MAINTENANCE UNDER SERVICE CONTRACT

ACTIONS TAKEN**PERFORMED PREVENTIVE MAINTENANCE**

-tested charging of the unit: good, tested also xray exposure: good
-checked and cleaned whole unit
-checked and cleaned boards and fuses
-checked 9volts, CMOS and 20pcs batteries: good
-checked batteries detection both (+) and (-) side: good
-checked voltage drop check mode: good
-checked mechanical movements, applied WD40
-checked DR system, checked FPD: good
-checked MWL and DICOM sending: good
-performed xray tube calibration: good ; gathered calibration data
-tested xray and image acquisition: good
-observed warning HDD notification, need to free up space
-endorsed to end user and biomed to delete old studies to free up space

REMARKS/RECOMMENDATIONS

UNIT OPERATIONAL

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 05, 2026	11:00AM	1:00PM	SALES INVOICE
			PURCHASE ORDER
			SC

SERVED BY:


DANIEL NOPPE
NAME & SIGNATURE

ACKNOWLEDGED BY:


JONIEL JANSEN TESARA
NAME & SIGNATURE

CONTACT NO: _____

EMAIL ADD: _____

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting