

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DAVAO DOCTORS HOSPITAL ADDRESS: DAVAO CITY
SERVICE REQUESTED BY: DEPARTMENT: XRAY ROOM #1
EQUIPMENT/MODEL: FLEXAVISION F3 SERIAL NO.: 41B4AC443001

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: SCHEDULED PREVENTIVE MAINTENANCE UNDER SERVICE CONTRACT

ACTIONS TAKEN**PERFORMED PREVENTIVE MAINTENANCE**

-checked and cleaned whole unit
-checked and cleaned generator cabinet
-checked and cleaned boards and fuses
-checked and cleaned, CH200M, bucky stand and ZS5DS
-adjusted vertical drop of CH200m
-checked mechanical movements, applied WD40
-unit uses only General Radiography mode / 2nd tube
-end user uses 701FPD and a separate DR system
-checked and cleaned system unit for CXDI NE
-checked MWL and DICOM sending: good
-performed xray tube calibration both tube 1 and tube 2 FVR only
-tested xray and image acquisition: good
-observed defective UPS of the CXDI NE PC, endorsed to biomed defective UPS
-endorsed to biomed and user to delete old studies to free up space
-exposure count = tube1=1450 , tube2=585,960

REMARKS/RECOMMENDATIONS

UNIT OPERATIONAL

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 05, 2026	3:00PM	6:00PM	SALES INVOICE
			PURCHASE ORDER
			SC

SERVED BY:


DANIEL NOPPE
NAME & SIGNATURE

ACKNOWLEDGED BY:


JONIEL JANSEN TESARA
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting