

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: CHONG HUA MEDICAL MALL ADDRESS: CEBU CITY, CEBU
SERVICE REQUESTED BY: BIOMED DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED PRO MC WITH CXDI (ROOM 1 & ROOM 2) SERIAL NO.:

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☒ OTHERS

COMPLAINT: NEW DICOM PRINTER INTEGRATION

ACTIONS TAKEN

-VISITED SITE
-CONFIGURED DICOM PRINTER TO CXDI PC FOR ROOM 1 AND ROOM 2
-TESTED PING AND C-ECHO FOR BOTH PC; SUCCESSFUL
-TESTED PRINTING OUTPUT FOR ROOM 1; SUCCESSFUL
-TESTED PRINTING OUTPUT FOR ROOM 2; SUCCESSFUL
-ENDORSED TO END USER HOW TO PRINT IMAGES
-ENDORSED UNIT TO END USER AND BIOMED

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION AND PRINTING CAN NOW BE USED.

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
APRIL 28, 2026	8:30 AM	9:45 AM	SALES INVOICE
			PURCHASE ORDER

SERVED BY:

JASPER B. REDOBLE
NAME & SIGNATURE

ACKNOWLEDGED BY:

SUNSHINE C. GAWAT, RRT/ JOHN PAUL S. SERMON
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting