

- ☐ Analytical  
☒ Medical  
☐ Testing Machine/Big Range  
☐ Balance

**TECHNICAL SERVICE REPORT**

CUSTOMER NAME: ST. VINCENT GENERAL HOSPITAL, INC. ADDRESS: CEBU CITY, CEBU  
SERVICE REQUESTED BY: SPC DEPARTMENT: RADIOLOGY  
EQUIPMENT/MODEL: SDR-100NW SERIAL NO.: LM8D9D4FC001

## SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☒ INSTALLATION ☐ OTHERS

## COMPLAINT:

## ACTIONS TAKEN

-VISITED SITE  
-DELIVERED 3 KVA UPS  
-CHECKED HOSPITAL OUTLET; 216.1 V AC  
-INSTALLED 3KVA UPS FOR SDR PC  
-CHECKED UPS OUTPUT VOLTAGE; 220.6 VAC  
-TESTED UPS OPERATION DURING NO POWER ON HOSPITAL; NORMAL  
-TESTED UNIT OPERATION; NORMAL  
-ENDORSED TEMPORARY DELIVERY RECEIPT TO END USER  
-ENDORSED UNIT TO END USER

## REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION AND READY TO USE.

## SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT  
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS  
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

## PARTS SUPPLIED/RECOMMENDED


| ITEM NO. | PART NUMBER | DESCRIPTION | QTY | DELIVERY RECEIPT |
|----------|-------------|-------------|-----|------------------|
| 1        |             | 3 KVA UPS   |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |

| DATE OF SERVICE | TIME STARTED | TIME FINISHED | BILLING STATEMENT |
|-----------------|--------------|---------------|-------------------|
| APRIL 29, 2026  | 8:30 AM      | 9:30 AM       | SALES INVOICE     |
|                 |              |               | PURCHASE ORDER    |

SERVICED BY:

  
JASPER B. REDOBLE  
NAME & SIGNATURE

ACKNOWLEDGED BY:

  
ARNEL SADUCAS, RRT  
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting