

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: CEBU NORTH GENERAL HOSPITAL ADDRESS: TALAMBAN, CEBU
SERVICE REQUESTED BY: END USER DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED PRO MF SERIAL NO.: 3M5262E56001

SERVICE CATEGORY

☐ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS EMERGENCY VISIT

COMPLAINT: X-RAY TUBE CONSOLE PANEL

ACTIONS TAKEN

-VISITED SITE
-TESTED X-RAY CONSOLE PANEL
-UPON EXTENDING THE ARM; X-RAY CONSOLE PANEL TURNS OFF
-UPON RETRACTING THE ARM; X-RAY CONSOLE PANEL TURNS ON
-CHEKED ALL CABLE CONNECTIONS; CABLES FROM ARM TO X-RAY CONSOLE PANEL IS EXTENDED TO ITS MAXIMUM LENGTH
-ADJUSTED CABLES SO THAT CABLE DISTANCE WILL JUST BE ENOUGH WHEN EXTENDED
-TESTED X-RAY CONSOLE OPERATION; STILL THE SAME
-CHECKED CABLES INSIDE THE X-RAY CONSOLE PANEL; ADJUSTED CABLES SO THAT IT WILL NOT EXTEND OR BREAK CAUSING IT TO LOSE CONNECTION TO X-RAY CONTROL CONSOLE PANEL
-ALSO NOTICED THAT CABLE CONNECTOR OF ROTATION OF X-RAY TUBE HAS A METAL PART EXPOSED; COVERED METAL PART WITH ELECTRICAL TAPE AS IT MIGHT CAUSE GROUNDING WHEN IN CONTACT TO METAL
-RESET CABLE CONNECTIONS OF X-RAY CONSOLE PANEL
-TESTED UNIT X-RAY UNIT OPERATION; NORMAL FUNCTION
-ENDORSED TO END USER

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION. OBSERVE ARM FUNCTION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
APRIL 29, 2026	11:30 AM	1:00 PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:

JASPER B. REDOBLE

NAME & SIGNATURE

ACKNOWLEDGED BY:

CLAROS, DIONISIO, RRT

NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting