

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ACE MT. SAMAT MEDICAL CENTER ADDRESS: PILAR, BATAAN
SERVICE REQUESTED BY: DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: MOBILE DART EVOLUTION MX8C SERIAL NO.: MQ00005D7004

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT:**ACTIONS TAKEN**

- TESTED UNIT BEFORE CONDUCTING PREVENTIVE MAINTENANCE = OK
- UPON POWER: U01, U02, U03 ALARM MESSAGES APPEAR (ALARMS ONLY, NOT AN ERROR)
- CLEANED UNIT
- CHECKED MECHANICAL MOVEMENTS AND EMERGENCY BRAKE MOVEMENT = OK
- CHECKED COLLIMATOR FUNCTIONS
- CHECKED CMOS BATTERY VOLTAGE = 3.052 V
- CHECKED MAIN BATTERY VOLTAGE = +130.3 V, -129.7 V
- CHECK EMERGENCY BRAKE BATTERIES = 9.36 V, 9.38 V
- TESTED DR SYSTEM = OK
- PERFORMED IMAGE ACQUISITION
- EXPOSURE COUNT = 3647 EXPOSURES
- RESET U01, U02, U03 ALARMS WITH CONSENT OF END-USER
- UNIT IS IN WORKING CONDITION

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

- ☒ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/06/2026	6 PM	8 PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:

ENGR. MARTEEN CRUZ

NAME & SIGNATURE

ACKNOWLEDGED BY:

ZERIAN RAY TUGUINAY, RRT

NAME & SIGNATURE

CONTACT NO:

09532966821

EMAIL ADD:

acentsamat.radiologydept@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting