

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ACE MT. SAMAT MEDICAL CENTER ADDRESS: PILAR, BATAAN
SERVICE REQUESTED BY: DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: OPESCOPE ACTENO FD SERIAL NO.: MQD6DBDD8001

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT:

ACTIONS TAKEN

- TESTED UNIT BEFORE CONDUCTING PREVENTIVE MAINTENANCE = OK
- CLEANED UNIT
- CHECKED MECHANICAL MOVEMENTS
- CHECKED BUTTON OPERATIONS
- CHECKED WHEEL MOVEMENTS
- CHECKED LOCK MECHANISMS
- CHECKED RADIOGRAPHY AND FLUOROSCOPY FUNCTION
- PERFORMED FVR AND FVF CALIBRATION
- PERFORMED FPD CALIBRATION
- PERFORMED TEST IMAGE ACQUISITION
- TUBE EXPOSURE COUNT = 5827 EXPOSURES
- UNIT IS IN WORKING CONDITION

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

- ☒ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/06/2026	8 PM	10:30 PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:

ENGR. MARTEEN CRUZ

NAME & SIGNATURE

ACKNOWLEDGED BY:

ZERIAN JAY TUGUINAY, RRT

NAME & SIGNATURE

CONTACT NO:

09532966821

EMAIL ADD:

acemtsamat.radiologydept@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting