

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: REGION II TRAUMA MEDICAL CENTER ADDRESS: BAYOMBONG MEDICAL CENTER
SERVICE REQUESTED BY: ROUJAHN M. DURANTE DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: FLEXAVISION HB SERIAL NO.: 4109C4D22004

SERVICE CATEGORY

☐ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☒ OTHERS DELIVERY

COMPLAINT:

ACTIONS TAKEN

- Delivered X-ray tube Serial Number: RM6D88AFB050 Model: 0.6/1.2 P324DK-85 Part Number (PN): 582-24486-50
- Performed cleaning of the X-ray tube mounting and accessories prior to complete dismantling.
- Dismantled the X-ray tube cooling fan cover, mounting, wiring, collimator, and high-voltage cable.
- Dismantled the old defective X-ray tube and installed a new X-ray tube.
- Installed the following X-ray tube accessories: low-tension wiring, high-voltage cable (with application of high-voltage grease), collimator, cover,
- Performed X-ray tube aging as a preparatory procedure prior to X-ray tube calibration, to ensure stable and reliable exposure output.
- Performed X-ray tube calibration for small and large focal spots to ensure accurate exposure performance.
- Checked and verified X-ray exposure output using a COBIA meter to ensure accurate dose performance.
- Checked the number of exposures prior to replacement; exposure count recorded at 320,118.
- Returned waste material of x-ray tube Model: 0.6/1.2P324DK-85 SN:RM65887001 YEAR 2018-07-03

REMARKS/RECOMMENDATIONS

Unit operational

SUBMITTED ORIGINAL DOCUMENTS

- ☐ STICKER ☒ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☒ CALIBRATION DATA (MEDICAL) ☒ INVOICE/BILLING ☒ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 07,2026	8:00AM	4:00PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:



DENNIS M.R.AZORES

NAME & SIGNATURE

ACKNOWLEDGED BY:



ROUJAHN M. DURANTE

NAME & SIGNATURE

CONTACT NO:

09395684009

EMAIL ADD:

radiology@riitmc.doh.gov.ph

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting