



SHIMADZU

Philippines Corporation

19F Marajo Tower, 26th St. cor. 4th Ave., BGC, Taguig City 1634 Metro Manila
(632) 8 869-9563 / (0917) 897-2057 (632) 8 519-9285 spcmanila@shimadzu.com.ph

No. 20260428-2-DA

- ☐ Analytical
- ☒ Medical
- ☐ Testing Machine/Big Range
- ☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ACE SAN JOSE DEL MONTE MOZUN MEDICAL CENTER ADDRESS: SAN JOSE BULACAN
SERVICE REQUESTED BY: BEA DEBUYAN DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED MC + CANON CXDI SERIAL NO.: MC927C4C1017

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT:

ACTIONS TAKEN

- Verified the input voltage (UV = 224 V, VW = 219 V, UW = 224 V) and confirmed output voltage readings of UV = 400 V, UW = 406 V, and VW =
- Verified low voltage parameters with measured values of L0 - L100 at 105.9 V and A0 - A200 at 203 V, confirming normal operation.
- Performed cleaning of the CH-200M ceiling support, table bucky, vertical bucky, and vacuum high voltage generator as part of preventive maintenance.
- Inspected the mechanical condition of the telescopic arm, table bucky, vertical bucky, and collimator and confirmed proper operation.
- Conducted X-ray tube aging preparation to ensure system stability before performing FVR calibration.
- Conducted FVR calibration to verify and ensure accurate and reliable X-ray exposure output performance.
- Verified the collimator alignment to ensure accurate alignment of the light field with the X-ray beam.
- Verified the total number of exposure counts, recorded at 38,162 exposures.
- Verified DICOM storage functionality and generated a sample DICOM image to confirm proper system operation.

REMARKS/RECOMMENDATIONS

UNIT OPERATIONAL

SUBMITTED ORIGINAL DOCUMENTS

- ☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
- ☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
- ☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
APRIL 28, 2026	1:00PM	4:00PM	SALES INVOICE
			PURCHASE ORDER

SERVED BY:

ENGR. DENNIS M. RAZORES/KEVIN GAROCHÉ
NAME & SIGNATURE

ACKNOWLEDGED BY:

ABIGAIL J. BATACANDOLO, RRT
NAME & SIGNATURE

CONTACT NO:

09171336865

EMAIL ADD:

radiologysjdmmmc@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting