



SHIMADZU

Philippines Corporation

19F Marajo Tower, 26th St. cor. 4th Ave., BGC, Taguig City 1634 Metro Manila
(632) 8 869-9563 / (0917) 897-2057 (632) 8 519-9285 spcmanila@shimadzu.com.ph

No. 20260430-3-DA

- ☐ Analytical
- ☒ Medical
- ☐ Testing Machine/Big Range
- ☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: PREMIERE MEDICAL CENTER ADDRESS: CABANATUAN CITY
SERVICE REQUESTED BY: ASSET MANAGEMENT DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: MOBILEART ECO MUX-10 SERIAL NO.: 3YCFC6BD1001

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT:

ACTIONS TAKEN

- Performed FVR calibration (Focus 1.2) at 40 kV - 125 kV and 80 mA - 160 mA to ensure reliable exposure output.
- Performed X-ray tube aging at 40 kVp - 120 kVp and 10 mAs in preparation for FVR calibration.
- Checked X-ray exposure output using a Raysafe meter to ensure patient safety.
- Checked the mechanical functions of the mobile X-ray unit, including tube angulation, tube rotation, and the telescopic arm.
- Cleaned the mobile X-ray unit and removed hair from the caster wheels.
- Performed collimator alignment checking to ensure proper alignment between the lamp light field and the X-ray beam.
-
-
-
-
-

REMARKS/RECOMMENDATIONS

UNIT OPERATIONAL

SUBMITTED ORIGINAL DOCUMENTS

- ☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
- ☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
- ☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
APRIL 30, 2026	10:00AM	12:00PM	SALES INVOICE
			PURCHASE ORDER

SERVED BY:

DENNIS MARTIN R. AZORES / Kevin Garoche
NAME & SIGNATURE

ACKNOWLEDGED BY:

ANALYN BALARIAVELLA YUTE
NAME & SIGNATURE

CONTACT NO:

09085006502

EMAIL ADD:

ymri2018@gmail.com / ella.yute@pmcne.ph

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting