

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DR. JOSE N. RODRIGUEZ MEMORIAL HOSPITAL AND SANITARIUM ADDRESS: CALOOCAN CITY
SERVICE REQUESTED BY: DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: FLEXAVISION F4 WITH CH-200M SERIAL NO.: MQC353CCC001

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☒ OTHERS

COMPLAINT:**ACTIONS TAKEN**

- PERFORMED NORMAL OPERATION CHECK BEFORE CONDUCTING PMS = OK
- CLEANED UNIT
- CHECKED MECHANICAL MOVEMENTS
- CHECKED COLLIMATOR FUNCTIONS
- PERFORMED FVR AND FVF CALIBRATION ON TUBE 1 FOR BOTH SMALL AND LARGE FOCUS
- PERFORMED FVR CALIBRATION ON TUBE 2 FOR SMALL FOCUS (LARGE FOCAL SPOT IS NOT WORKING)
- PERFORMED IMAGE QUALITY ASSESSMENT ON DR SOFTWARE; OBSERVED IMAGE CUTOFF ON HEAD SIDE (FOR 1ST TUBE)
- ADJUSTED COLLIMATOR FOR IMAGE CUTOFF
- PERFORMED FIELD OF VIEW ADJUSTMENT FOR TUBE 1
- PERFORMED CALIBRATION FOR WIRELESS FLAT PANEL DETECTOR
- CHECKED ALL CONSOLE/S FUNCTIONS = OK
- PERFORMED SYSTEM BACKUP
- SID LOCK TO BE ADJUSTED FOR 40 IN.: DUE TO LACK OF RESOURCE TO ACCESS RAIL, ADJUSTMENT WILL BE TO FOLLOW
- COLLIMATOR FOR TUBE 1 (FLUOROSCOPY WITH TABLE) IS ROTATING, LOCKED IN PLACE TO MAINTAIN ALIGNMENT ON IMAGE
- TOTAL EXPOSURE COUNT: TUBE 1 (R/F) = 17679 EXPOSURES, TUBE 2 (CEILING MOUNTED) = 122807 EXPOSURES
- UNIT IS IN WORKING CONDITION
- NOTE: REMOTE CONSOLE TABLE KEYBOARD SLIDE IS FAULTY; OBSERVE CAUTION WHEN PULLING

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☒ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
04/29/2026	1 PM	7:30 PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:

ENGR. MARTEEN CRUZ

NAME & SIGNATURE

ACKNOWLEDGED BY:

JOSELITO D. CRISTO JR.

NAME & SIGNATURE

CONTACT NO:

09301211005

EMAIL ADD:

guzman@gmail.com/djnmhsradiologyradtec

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting