

- ☐ Analytical  
☒ Medical  
☐ Testing Machine/Big Range  
☐ Balance

**TECHNICAL SERVICE REPORT**

CUSTOMER NAME: DAVAO REGIONAL MEDICAL CENTER ADDRESS: APOKON, TAGUM CITY  
SERVICE REQUESTED BY: DEPARTMENT: OPD-XRAY  
EQUIPMENT/MODEL: RAD SYSTEM / RADSPED FIT DR SERIAL NO.: MP41F0BF2004

## SERVICE CATEGORY

☒ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: LOOSE BEARING AT XRAY TUBE

## ACTIONS TAKEN

-CHECKED THE UNIT  
-ASKED END-USER FOR CONCERNS, OBSERVED LOOSE BEARING FOR TUBE ROTATION  
-CHECKED THE BEARING, OK  
-TIGHTENED LOOSE SCREW FOR THE BEARING  
-CHECKED TUBE ROTATION AND APPLIED LUBRICANT, OK  
-TESTED OTHER MOVEMENTS, OK  
-TESTED IMAGE ACQUISITION, OK  
-ENDORSED TO END-USER  
-DELIVERED 1PC PATIENT SLIDE AND 2PCS LEAD GOGGLES  
-DELIVERY RECEIPT SIGNED BY END USER, BIOMED AND MMS

## REMARKS/RECOMMENDATIONS

-UNIT IN GOOD WORKING CONDITION

## SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT  
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS  
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

## PARTS SUPPLIED/RECOMMENDED

| ITEM NO. | PART NUMBER | DESCRIPTION | QTY | DELIVERY RECEIPT |
|----------|-------------|-------------|-----|------------------|
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |

| DATE OF SERVICE | TIME STARTED | TIME FINISHED | BILLING STATEMENT | SALES INVOICE | PURCHASE ORDER | WARRANTY |
|-----------------|--------------|---------------|-------------------|---------------|----------------|----------|
| 04/30/2026      | 10:00AM      | 11:30AM       |                   |               |                |          |
|                 |              |               |                   |               |                |          |
|                 |              |               |                   |               |                |          |

SERVICED BY:

JOSEPH REY B. DUMASIG  
NAME & SIGNATURE

ACKNOWLEDGED BY:

CRISANTO P. TORCULAS RRT  
NAME & SIGNATURE

CONTACT NO:

09564071659

EMAIL ADD:

erisantotorcullas@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting