

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ACE MEDICAL CENTER – LEGAZPI ADDRESS: LEGAZPI CITY
SERVICE REQUESTED BY: DEPARTMENT: CATHETERIZATION LABORATORY
EQUIPMENT/MODEL: TRINIAS UNITY C16 SERIAL NO.: M1EFD34C9001

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: PREVENTIVE MAINTENANCE OF TRINIAS UNITY C16

ACTIONS TAKEN

1. CLEANED AND CHECKED MECHANICAL MOVEMENTS OF THE SYSTEM
2. RE-GREASED HIGH VOLTAGE CABLES, XRAY TUBE, HIGH VOLTAGE CABINET AND STARTER CABINET SIDES
3. PERFORMED TUBE WARM-UP
4. PERFORMED FLASH CALIBRATION
5. PERFORMED FVR CALIBRATION, SMALL AND LARGE FOCUS
6. PERFORMED FVF CALIBRATION DC MODE
7. PERFORMED FPD CALIBRATION, GAIN AND SPECIFIC MODES
8. RE-TIGHTENED BOLTS OF FIXED RAILS
9. PERFORMED FLUOROSCOPY AND RADIOGRAPHY EXPOSURE TESTS
10. PERFORMED 3D RADIOGRAPHY EXPOSURE TEST
11. TRANSMITTED SAMPLE IMAGES TO VRS STATION; GOOD
12. CHECKED HEMODYNAMICS; GOOD
13. CHECKED FPD AND HEAT EXCHANGER COOLANT LIQUID; GOOD
14. ENDORSED THE SYSTEM TO BIOMED

REMARKS/RECOMMENDATIONS 1. DONE PREVENTIVE MAINTENANCE OF TRINIAS UNITY C16
2. SYSTEM IS WORKING. ADVISED BIOMED TO INFORM END-USER NOT TO USE THE SYSTEM BECAUSE THE UPS IS NOT WORKING PROPERLY

SUBMITTED ORIGINAL DOCUMENTS

☒ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 08, 2026	4PM	10PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:

ACKNOWLEDGED BY:

ENGR. JOHN MICHAEL MARFIL
NAME & SIGNATURE

BEN N. VILLAFLORES
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD: villaflores704@gmail.com