

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: BICOL REGIONAL HOSPITAL AND MEDICAL CENTER ADDRESS: LEGAZPI CITY
SERVICE REQUESTED BY: _____ DEPARTMENT: RADIOLOGY DEPARTMENT
EQUIPMENT/MODEL: MOBILEDART MX7 SERIAL NO.: MPF16817C007

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS _____

COMPLAINT: PREVENTIVE MAINTENANCE OF MOBILEDART MX7

ACTIONS TAKEN

1. CLEANED AND CHECKED MECHANICAL MOVEMENTS OF THE SYSTEM
2. PERFORMED TWO-POINT AND POINT-TO-POINT TUBE CURRENT ADJUSTMENTS OF SMALL AND LARGE FOCUS
3. CHECKED EXPOSURE COUNT; 10327
4. CHECKED DR SYSTEM AND WIRELESS DETECTOR CONNECTION; GOOD
5. PERFORMED RADIOGRAPHY EXPOSURE TESTS; GOOD
6. CHECKED STORAGE AND MWL IN CONNECTION SETTING; PING AND ECHO ARE SUCCESSFUL
7. TRANSMITTED SAMPLE IMAGES TO STORAGE; GOOD
8. ENDORSED THE SYSTEM TO END-USER

REMARKS/RECOMMENDATIONS

1. DONE PREVENTIVE MAINTENANCE OF MOBILEDART MX7
2. SYSTEM IS WORKING AND READY TO USE

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED


ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED
MAY 09, 2026	4PM	5PM

BILLING STATEMENT _____
SALES INVOICE _____
PURCHASE ORDER _____

SERVICED BY:

ACKNOWLEDGED BY:


ENGR. JOHN MICHAEL MARFIL
NAME & SIGNATURE


CHRISTIAN ANGELO AMACIO
NAME & SIGNATURE

CONTACT NO: _____
EMAIL ADD: _____