

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ZONE MEDICAL AND INTERVENTION HOSPITAL, INC ADDRESS: LIGAO CITY, ALBAY
SERVICE REQUESTED BY: _____ DEPARTMENT: OPERATING ROOM
EQUIPMENT/MODEL: OPESCOPE ACTENO FD TYPE SERIAL NO.: MQD6DBDF5006

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS _____

COMPLAINT: PREVENTIVE MAINTENANCE OF OPESCOPE ACTENO FD TYPE

ACTIONS TAKEN

1. CLEANED AND CHECKED MECHANICAL MOVEMENTS OF THE SYSTEM
2. CHECKED SYSTEM OPERATION PRIOR PMS; GOOD
3. CHECKED MEC STATUS AND ENABLED
4. PERFORMED FVR AND FVF (DC AND PULSE FLUOROSCOPY); SYSTEM ENCOUNTERED ERROR RELATING TO CONNECTION OF FXIO ON THE SYSTEM, DRIVER MIGHT BE CORRUPTED OR MISSING
5. DISABLED MEC AND CHECKED FXIO DRIVER; FXIO IS NOW WORKING PROPERLY
6. PERFORMED WHITELIST; TO INCLUDE FXIO DRIVER FOR THE ACCEPTED SOFTWARES IN THE SYSTEM
7. ENABLED MEC AND CHECKED FXIO CONNECTION; NO ISSUES OBSERVED
8. PERFORMED FPD CALIBRATION
9. PERFORMED RADIOGRAPHY AND FLUOROSCOPY EXPOSURE TESTS
10. CHECKED EXPOSURE COUNT; 1429
11. ENDORSED THE SYSTEM TO END-USER

NOTE: HOSPITAL ENCOUNTERED POWER INTERRUPTION MULTIPLE TIMES, RECOMMENDED TO HAVE 6KVA ONLINE UPS FOR THE SYSTEM

REMARKS/RECOMMENDATIONS

1. DONE PREVENTIVE MAINTENANCE OF OPESCOPE ACTENO FD
2. SYSTEM IS WORKING AND READY TO USE; RECOMMENDED TO HAVE 6KVA ONLINE UPS

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED
MAY 12, 2026	11AM	3PM

BILLING STATEMENT _____
SALES INVOICE _____
PURCHASE ORDER _____

SERVICED BY:

ACKNOWLEDGED BY:


ENGR. JOHN MICHAEL MARFIL
NAME & SIGNATURE


MARIA EVAREY ADONES
NAME & SIGNATURE

CONTACT NO: _____

EMAIL ADD: zmihrad001@gmail.comss