

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ANTIQUE MEDICAL CENTER ADDRESS: SAN JOSE, ANTIQUE
SERVICE REQUESTED BY: END-USER DEPARTMENT: O.R
EQUIPMENT/MODEL: MOBILE SURGICAL C-ARM OPESCOPE ACTENO FD SERIAL NO.: MQD6DBDE5002

SERVICE CATEGORY

☒ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS _____

COMPLAINT: PREVENTIVE MAINTENENCE SERVICE WARRANTY (2ND VISIT OF 1ST YEAR) 2/4

ACTIONS TAKEN

-PERFORMED GENERAL CLEANING
-CHECKED HOSPITAL POWER OUTLET; 219.7VAC
-CHECKED AVR POWER SUPPLY OUTPUT; 219.9VAC
-CHECKED ALL FUSES; OK
-PERFORMED FVF AND FVR CALIBRATION
-CHECKED TUBE EXPOSURE COUNT; 3,130
-CHECKED ERROR LOG
-PERFORMED FPD CALIBRATION
-CHECKED FUNCTION AND OPERATION; OK
-TESTED IMAGE AQUISITION
-ATTACHED PMS STICKER

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING GOOD CONDITION

SUBMITTED ORIGINAL DOCUMENTS


☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
			SALES INVOICE
			PURCHASE ORDER

SERVED BY:


JOMARIE T. PENAFLO
NAME & SIGNATURE

ACKNOWLEDGED BY:


ERAIZA MAE DAVID
NAME & SIGNATURE

CONTACT NO:**EMAIL ADD:**

amci.did25@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting