

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: OUR LADY OF PEACE HOSPITAL ADDRESS: PARAÑAQUE CITY
SERVICE REQUESTED BY: DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED FIT SERIAL NO.: MP41F0BFA001

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☒ INSTALLATION ☐ OTHERS

COMPLAINT:**ACTIONS TAKEN**

- FLOOR WORKS BY ENGINEERING IS DONE
- REPOSITIONED X-RAY RADIOGRAPHY TABLE AND BUCKY STAND
- LEVELED TABLE AND BUCKY STAND WITH FLOOR
- FIXED X-RAY RADIOGRAPHY TABLE POSITION
- ALIGNED X-RAY TUBE WITH BUCKY STAND ON 40 INCHES AND 72 INCHES DISTANCE
- FIXED BUCKY STAND POSITION
- RECONNECTED POWER CABLE WITH CONDUIT
- CONNECTED BUCKY STAND WIRES TO GENERATOR
- ORGANIZED WIRINGS ON GENERATOR SIDE
- PERFORMED IMAGE ACQUISITION
- CONDUCTED USER TRAINING AS PER REQUEST OF END-USERS
- UNIT IS IN WORKING CONDITION

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

- ☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/11/2026	1 PM	8 PM	SALES INVOICE
			PURCHASE ORDER

SERVED BY:

ENGR. MARTEEN CRUZ/KEVIN GAROCHE
NAME & SIGNATURE

ACKNOWLEDGED BY:

SHENNAF WHOSE T. NUIQUE
NAME & SIGNATURE

CONTACT NO:**EMAIL ADD:**

09762696618

radiology.olph2021@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting