

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: DON JOSE S. MONFORT MEDICAL CENTER ADDRESS: BAROTAC NUEVO, ILOILO
SERVICE REQUESTED BY: END USER DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: RADSPED PRO MC W/ CANON DR SERIAL NO.: MPC5C169B010

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☒ OTHERS P.O.

COMPLAINT: DELIVERED 2PCS BATTERY PACK LB-4A FOR CANON FPD (NCS)

ACTIONS TAKEN

-DELIVERED 2PCS BATTERY PACK LB-4A
-INSTALLED 1 BATTERY PACK TO THE CANON FPD
-THE END USER WAS INSTRUCTED ON THE PROPER USE AND PROPER CHARGING PROCEDURE OF THE NEW BATTERY
-TESTED BATTERY CHARGING; OK
-TESTED IMAGE AQUISITION; OK
-ENDORSED TO THE END USER

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING GOOD CONDITION / INVOICE AND DELIVERY RECEIPT WILL BE FOLLOWED

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 11, 2026	3:40 P.M	4:10 P.M	SALES INVOICE
			PURCHASE ORDER

SERVED BY:

JOMARIE T. PENAFLO

NAME & SIGNATURE

ACKNOWLEDGED BY:

KIMBER WILLIAM CION

NAME & SIGNATURE

CONTACT NO:**EMAIL ADD:**

cionkimberwilliam@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting