

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: ST. PAUL'S HOSPITAL OF ILOILO INC. ADDRESS: ILOILO CITY
SERVICE REQUESTED BY: END-USER DEPARTMENT: X-RAY 1
EQUIPMENT/MODEL: R/F SYSTEM/FLEXAVISION F4 W/ CH-200M SERIAL NO.: MQC353CF1004

SERVICE CATEGORY

☒ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: UNSUAL SOUND IS COMING FROM THE SPEAKER INSIDE THE EXAMINATION ROOM

ACTIONS TAKEN

-CHECKED THE SPEAKER WIRE FOR ANY OPEN WIRE OR POSSIBLE RAT BITES; NO OPEN WIRE
-CHECKED CONTROL KNOB OF THE MICROPHONE; OK
-CHECKED CONTROL KNOB FOR SPEAKER VOLUME; OK
-THE UNSUAL SOUND CAN BE HEARD WHEN THE SPEAKER VOLUME IN THE EXAMINATION ROOM IS SET TO FULL VOLUME.
-SET THE SPEAKER VOLUME TO MINIMUM TO PREVENT THE OCCURRENCE OF UNSUAL SOUND.
-TESTED THE UNIT

REMARKS/RECOMMENDATIONS

UNIT IS WORKING GOOD CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
MAY 9, 2026	3:00 P.M	3:30 P.M	SALES INVOICE
			PURCHASE ORDER
			UNDER WARRANTY

SERVED BY:


JOMARIE T. PENAFLO
NAME & SIGNATURE

ACKNOWLEDGED BY:


VINCENT D. GEOLAMIN
NAME & SIGNATURE

CONTACT NO:

09623733522

EMAIL ADD:

ysocuelan@gmail.com / sphdidct@yahoo.co

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting