

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: SOUTHERN PHILIPPINES MEDICAL CENTER ADDRESS: DAVAO CITY
SERVICE REQUESTED BY: DEPARTMENT: CATHLAB ROOM
EQUIPMENT/MODEL: CVS / TRINIAS UNITY SMART EDITION C16 SERIAL NO.: M1F03D8DA001

SERVICE CATEGORY

☐ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: CHECKUP OF THE HEMODYNAMICS

ACTIONS TAKEN

CHECKED THE UNIT, NO ERROR ENCOUNTERED
ASKED END USER FOR CONCERNS, WINDOWS UPDATE CONCERN POP OUT
CHECKED GROUP POLICY EDITOR
CONFIGURED AUTOMATIC UPDATES TO DISABLED
CHECKED WINDOWS UPDATE AND WINDOWS VERSION, OK
CHECKED NETWORK SETTINGS, INTERNET ACCESS OBSERVED
TURNED OFF INTERNET ACCESS
ADVISED END USER TO USE THE HEMODYNAMICS FOR HEMODYNAMICS ONLY
CHECKED LANGUAGE SETTINGS, ALREADY SET TO ENGLISH
CHECKED FUNCTIONALITY, OK
ENDORSED TO END USER

REMARKS/RECOMMENDATIONS

-UNIT IN GOOD WORKING CONDITION

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT	SALES INVOICE	PURCHASE ORDER	WARRANTY
05/15/2026	9:30AM	11:00PM				

SERVICED BY:

JOSEPH REY B. DUMASIG
NAME & SIGNATURE

ACKNOWLEDGED BY:

MONETH MADULIN
NAME & SIGNATURE

CONTACT NO:

09052675337

EMAIL ADD:

mmdago0725@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting