

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: CARAGA REGIONAL HOSPITAL ADDRESS: SURIGAO CITY
SERVICE REQUESTED BY: DEPARTMENT: OPERATING ROOM
EQUIPMENT/MODEL: OPESCOPE ACTENO I.I. WITH SDR SERIAL NO.: MPD6D8291006

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☒ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT:

ACTIONS TAKEN

SCHEDULED PREVENTIVE MAINTENANCE OF OPESCOPE ACTENO UNDER SERVICE CONTRACT - 1ST YEAR, 2ND VISIT

- Performed test fluoroscopy and radiography exposure, unit working normally with no errors encountered.
- Checked power cable plug, tightened some screws, also checked the monitor cart cable, pins are in good shape.
- Checked mechanical movements, observed slide motion can be moved even when the c-arm is locked, checked and the brake pads are worn, arm hold assy recommended for replacement
- Checked monitor card and C-arm mobile boards and fuses; Good
- Checked and cleaned SDR-120 (PC) of the monitor cart
- Accessed maintenance tool of the monitor cart, checked event and error logs
- Checked the date and time settings of the PC and WhaCont and synchronized both systems.
- Performed FVR and FVF calibration of the tube
- Verified overall tube exposure = 16,398 total exposures
- Tested spot radiography and fluoroscopy exposure, Good
- Tested Collimation movements and Fluoroscopy save, Good

- Endorsed to end-user

REMARKS/RECOMMENDATIONS

UNIT IN GOOD WORKING CONDITION

ARM HOLD ASSY FOR REPLACEMENT

SUBMITTED ORIGINAL DOCUMENTS

- ☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☐ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED


ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT
1		ARM HOLD ASSY - For Replacement		

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/12/2026	2:00PM	5:00PM	SALES INVOICE
			PURCHASE ORDER
			AS PER S.C.

SERVICED BY:


D. NOPRE
NAME & SIGNATURE

ACKNOWLEDGED BY:


LANDENDEEFER S. RIVAS, RRT
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

crhradiologyfebruary2021@gmail.com

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting