

- ☐ Analytical
☒ Medical
☐ Testing Machine/Big Range
☐ Balance

TECHNICAL SERVICE REPORT

CUSTOMER NAME: RIZAL MEDICAL CENTER ADDRESS: PASIG CITY
SERVICE REQUESTED BY: DEPARTMENT: RADIOLOGY
EQUIPMENT/MODEL: FLEXAVISION F3 WITH CH-200M SERIAL NO.: MPEA63CA1001

SERVICE CATEGORY

☐ WARRANTY ☐ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☒ INSTALLATION ☒ OTHERS

COMPLAINT:**ACTIONS TAKEN**

- DISMANTLED OLD X-RAY TUBE = MODEL: 0.6/1.2P324DK-85; SN: RM6D858A1007
- DELIVERED AND INSTALLED NEW X-RAY TUBE = MODEL: 0.6/1.2P324DK-85; SN: RM6D88AG3031
- CHECKED AND CONFIGURED TUBE SETTINGS ON SYSTEM
- RESET TUBE EXPOSURE COUNT
- PERFORMED TUBE AGING
- PERFORMED FVR CALIBRATION FOR SMALL AND LARGE FOCUS
- COLLECTED CALIBRATION DATA
- UNIT IS IN WORKING CONDITION
- CHECKED WIRELESS DETECTOR SETTINGS FOR PERFORMING TEST IMAGE (CXDI-702C)
- CISCO ACCESS POINT CANNOT BE DETECTED ON DESKTOP SWITCH AND LAN: UPON TESTING, CONSOLE PORT OF CISCO ACCESS POINT IS DEFECTIVE
- TESTED WIRELESS DETECTOR CONNECTION (CXDI-702C) USING ARUBA ACCESS POINT OF UNOPERATIONAL MOBILE X-RAY SYSTEM (MX8 SHIMADZU): DETECTOR AND ACCESS POINT WAS ABLE TO COMMUNICATE BUT IT IS NOT COMPATIBLE WITH THE SYSTEM THEREFORE TEST IMAGE CANNOT BE PERFORMED
- ADJUSTED CLICK DOG FOR CENTERING OF X-RAY TUBE TO BUCKY STAND (TRANVERSAL DIRECTION)
- CHECKED BUCKY STAND: REPLACED MISSING SCREWS ON CASSETTE LOCK
- ENDORSED OLD X-RAY TUBE TO BIOMED. BIOMED

REMARKS/RECOMMENDATIONS

UNIT IS IN WORKING CONDITION. CISCO ACCESS POINT IS DEFECTIVE AND IS RECOMMENDED FOR REPLACEMENT. IT IS ALSO RECOMMENDED TO REPLACE DETECTOR THAT IS COMPATIBLE WITH THE SYSTEM.

SUBMITTED ORIGINAL DOCUMENTS

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

PARTS SUPPLIED/RECOMMENDED

ITEM NO.	PART NUMBER	DESCRIPTION	QTY	DELIVERY RECEIPT

DATE OF SERVICE	TIME STARTED	TIME FINISHED	BILLING STATEMENT
05/05/2026	7:30 AM	6:30 PM	SALES INVOICE
			PURCHASE ORDER

SERVICED BY:

ENGR. MARTEEN CRUZ/ENGR. RENE ROBIN CANO

NAME & SIGNATURE

ACKNOWLEDGED BY:

ANTHONY JOE ORATE

NAME & SIGNATURE

CONTACT NO:

09108637750

EMAIL ADD:

gov.ph/stephenbryan.1621@gmail.com/darw

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting