

- ☐ Analytical  
☒ Medical  
☐ Testing Machine/Big Range  
☐ Balance

**TECHNICAL SERVICE REPORT**

CUSTOMER NAME: Pasig City General Hospital ADDRESS: Pasig City  
SERVICE REQUESTED BY: End User DEPARTMENT: Radiology  
EQUIPMENT/MODEL: RADspeed Pro SERIAL NO.: MPC5C1173001

**SERVICE CATEGORY**

☐ WARRANTY ☒ CHECKUP/REPAIR ☐ PREVENTIVE MAINTENANCE ☐ INSTALLATION ☐ OTHERS

COMPLAINT: X-ray Tube Very Loud Noise, Exposure produced was very grainy, Radio current over

**ACTIONS TAKEN**

> Checked machine current condition  
> Large focus is busted (Filament not Ready Error), small focus will have a ready status indicator.  
> Upon test exposure, no loud sound encountered but error occurs (Radio Current Over). Sometimes able to expose without error.  
> Tested using super maintenance mode (bypass error detection), no error but time cuts off automatically at 0.3s or automatic cutoff at 26ms which indicates abnormality.  
> Checked resistance values for small and large focus. For small 0.5 to 1ohm, for large focus OL.  
> Checked FVR values (Filament Voltage Radiography), tried changing values as this could affect as directly stated from service manual  
> Error still occurs. Radio current over and time cuts off at 0.3ms or sometimes at 26ms  
> Checked overall tube exposure count: 176,850 total exposures  
> Regreased X-ray Tube  
> Checked stator resistance (P Type) : X-Y= 7.6ohms ok, X-Z = 34.7 ohms not ok  
> Checked measdsp, TmA values measured reached 544mA to 1996mA which prompts the machine to cut exposure immediately (0.3ms)  
> Verified to end user about x-ray tube's status during last times of usage:  
- scary loud noise (like a power drill sound)  
- X-ray images taken were grainy,  
- Chest X-rays and L-spine images taken were indistinguishable. (Grainy and lacks contrast)

**REMARKS/RECOMMENDATIONS**

X-ray tube is recommended for replacement. Unit cannot expose properly, recommended not to expose.

**SUBMITTED ORIGINAL DOCUMENTS**

☐ STICKER ☐ CALIBRATION CERTIFICATE ☐ PM CERTIFICATE ☐ COMPLETION REPORT  
☐ SYSTEM CHECK ☐ IQ DOCUMENTS ☐ OQ DOCUMENTS ☐ PM DOCUMENTS  
☒ SERVICE REPORT ☐ CALIBRATION DATA (MEDICAL) ☐ INVOICE/BILLING ☐ DELIVERY RECEIPT

**PARTS SUPPLIED/RECOMMENDED**

| ITEM NO. | PART NUMBER | DESCRIPTION | QTY | DELIVERY RECEIPT |
|----------|-------------|-------------|-----|------------------|
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |
|          |             |             |     |                  |

| DATE OF SERVICE | TIME STARTED | TIME FINISHED | BILLING STATEMENT |
|-----------------|--------------|---------------|-------------------|
| 5/5/2026        |              |               | SALES INVOICE     |
|                 |              |               | PURCHASE ORDER    |

**SERVED BY:**

Engr. Allan Mamangan  
NAME & SIGNATURE

**ACKNOWLEDGED BY:**

ALLAN PADRIGANO  
NAME & SIGNATURE

CONTACT NO:

EMAIL ADD:

WHITE: Customer Accounting

PINK: SPC Service Dept

GREEN: Customer End User

YELLOW: SPC Accounting