ChemQC Lot Report

Lot Information

Lot Number:	20251103J
Analysis Date:	2025-12-12

Test Summary

Total Channels Analyzed:	6
Channels Passed:	0
Channels Failed:	6
Final Lot Disposition:	**FAIL**

Lot Summary Result

Ink	Green Channel (G)	Red Channel (R)
51.9	**FAIL**	**FAIL**
33.39	**FAIL**	**FAIL**
51.9 LD	**FAIL**	**FAIL**

Report Generation

Operator:	Raphael Pham
Timestamp:	2026-01-05 13:34:56

Manual Verification

Operator Initials: _____

Date: _____

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.

51.9 Analysis

Lot: 20251103J

Green Channel Intensity **FAIL**

- SO2 Sensors: 0 passed, 6 failed (Total: 6)

Results Summary

Parameter	Specification	Value	Result	Batch Specification	Batch Value	Batch Result
SO2	> 10%	N/A	N/A	N/A	N/A	N/A
Performance	95% of wells pass	56.25% 52.0833% 42.7083% 30.2083% 33.3333% 41.6667%	FAIL FAIL FAIL FAIL FAIL FAIL	80% of sensors	10.0%	**FAIL**
Uniformity	1% (CV of Gci)	0.996% 0.7499% 0.7691% 0.7457% 0.8364% 0.7868%	PASS PASS PASS PASS PASS PASS	80% of sensors	100.0%	**PASS**

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.

Red Channel Intensity **FAIL**

- SO2 Sensors: 0 passed, 6 failed (Total: 6)

Results Summary

Parameter	Specification	Value	Result	Batch Specification	Batch Value	Batch Result
SO2	> 5.46%	N/A	N/A	N/A	N/A	N/A
Performance	95% of wells pass	56.25% 52.0833% 42.7083% 30.2083% 33.3333% 41.6667%	FAIL FAIL FAIL FAIL FAIL FAIL	80% of sensors	10.0%	**FAIL**
Uniformity	1% (CV of Gci)	0.5486% 0.772% 0.4992% 0.4564% 0.5366% 0.7064%	PASS PASS PASS PASS PASS PASS	80% of sensors	100.0%	**PASS**

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.

33.39 Analysis

Lot: 20251103J

Green Channel Intensity **FAIL**

- NH3 Sensors: 0 passed, 4 failed (Total: 4)

Results Summary

Parameter	Specification	Value	Result	Batch Specification	Batch Value	Batch Result
NH3	< -20%	N/A	N/A	N/A	N/A	N/A
Performance	95% of wells pass	93.75% 94.7917% 92.7083% 95.8333%	FAIL FAIL FAIL PASS	80% of sensors	10.0%	**FAIL**
Uniformity	1% (CV of Gci)	0.3472% 0.3734% 0.3713% 0.3917%	PASS PASS PASS PASS	80% of sensors	100.0%	**PASS**

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.

Red Channel Intensity **FAIL**

- NH3 Sensors: 0 passed, 4 failed (Total: 4)

Results Summary

Parameter	Specification	Value	Result	Batch Specification	Batch Value	Batch Result
NH3	< -20%	N/A	N/A	N/A	N/A	N/A
Performance	95% of wells pass	93.75% 94.7917% 92.7083% 95.8333%	FAIL FAIL FAIL PASS	80% of sensors	10.0%	**FAIL**
Uniformity	1% (CV of Gci)	0.349% 0.3359% 0.3501% 0.401%	PASS PASS PASS PASS	80% of sensors	100.0%	**PASS**

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.

51.9 LD Analysis

Lot: 20251103J

Green Channel Intensity **FAIL**

- SO2 Sensors: 0 passed, 6 failed (Total: 6)

Results Summary

Parameter	Specification	Value	Result	Batch Specification	Batch Value	Batch Result
SO2	> 10%	N/A	N/A	N/A	N/A	N/A
Performance	95% of wells pass	63.6364% 57.1429% 57.1429% 50.6494% 54.5455% 53.2468%	FAIL FAIL FAIL FAIL FAIL FAIL	80% of sensors	10.0%	**FAIL**

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.

Red Channel Intensity **FAIL**

- SO2 Sensors: 0 passed, 6 failed (Total: 6)

Results Summary

Parameter	Specification	Value	Result	Batch Specification	Batch Value	Batch Result
SO2	> 5.49%	N/A	N/A	N/A	N/A	N/A
Performance	95% of wells pass	63.6364% 57.1429% 57.1429% 50.6494% 54.5455% 53.2468%	FAIL FAIL FAIL FAIL FAIL FAIL	80% of sensors	10.0%	**FAIL**

CONFIDENTIAL: This document contains proprietary information and is intended for internal use only. Unauthorized disclosure or distribution is prohibited. This report is for Quality Control purposes only. R&D data is excluded.