SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Test laboratory (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

TEST REPORT

Basic safety test report: ECH60S handle

Test report IEC 60601-1 (edition 3.2) with EN 60601-1:2006+A13:2024 deviations, internally powered ME equipment · EN 60601-1:2006+A13:2024

Report number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECHELON 3000 60 mm Standard stapler, ECH60S, with primary lithium battery pack
Key data	Internally powered · type CF applied part · IPX0 · 40 W (public: IFU)
Scope	Example scope statement. Basic safety and essential performance of the battery-powered handle. Reloads are non-electrical and out of scope.

Classification (example)

Power source	Internally powered (primary lithium pack, user installed — public)
Applied part	Type CF (public: IFU symbol glossary)
Ingress protection and power	IPX0, 40 W (public: IFU)
Mode of operation	Intermittent (firing cycles) — example
Essential performance	No knife movement without activation of the firing trigger (public: IFU EMC section)

Clauses (example verdicts)

Clause	Requirement	Verdict (example)
4	Risk management process applied	Pass
8	Protection against electrical hazards	Pass
9	Mechanical hazards	Pass
11	Excessive temperatures (handle, battery)	Pass
13	Hazardous situations and fault conditions	Pass
14	Programmable electrical medical systems	Pass (with IEC 62304 file)
15.4.3	Batteries	Pass

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN