SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Biocompatibility laboratory (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

BIOLOGICAL EVALUATION

Biological evaluation report: staples and reload

Biological evaluation within a risk management process (EN ISO 10993-1:2025), with toxicological risk assessment (EN ISO 10993-17) · EN ISO 10993-1:2025

Report number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	Implantable staples and patient-contacting reload surfaces, GST60 family
Key data	Implant · tissue/bone · long-term (> 30 days) — categorisation example
Scope	Example scope statement. Evaluation of biological safety of the staples (implanted) and of transient contact surfaces of the reload. Handle surfaces evaluated separately.

Categorisation

Nature of contact	Implant device, tissue (staples); transient tissue contact (cartridge surfaces)
Duration	Long-term (staples); transient (cartridge)
Materials	Staple alloy: titanium alloy (not named publicly for GST; example); cobalt declared in EUDAMED

Endpoints and evidence (example)

Endpoint	Standard	Evidence (example)	Outcome (example)
Chemical characterisation	ISO 10993-18	Extractables, exhaustive extraction	Done
Toxicological risk assessment	ISO 10993-17	MoS for cobalt and alloy elements	MoS > 1
Cytotoxicity	ISO 10993-5	MEM elution	Non-cytotoxic
Sensitisation	ISO 10993-10	GPMT or rationale	Not sensitising
Irritation	ISO 10993-23	In vitro RhE	Non-irritant
Implantation / local effects	ISO 10993-6	Literature + predicate	Acceptable
Genotoxicity, systemic toxicity	ISO 10993-3/-11	Risk-based rationale	Acceptable

Conclusion (example)

The biological risks are acceptable for the intended use. The 2025 edition was applied; the gap assessment is linked to change event E4. Example conclusion only.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN