SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Manufacturer (example) (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

CLINICAL EVALUATION

Clinical evaluation report: GST60 reloads with ECHELON 3000

Clinical evaluation report under MDR Article 61 and Annex XIV Part A · MDR Art. 61, Annex XIV Part A, MDCG 2020-5, 2020-6, 2020-13 (example)

Report number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECH60S handle with GST60W/B/D/G/T reloads
Key data	Indications: transection, resection, anastomosis (public intended use)
Scope	Example scope statement. Clinical safety and performance across bariatric, colorectal and thoracic use, based on literature, predicate data and PMS. Colorectal PMCF planned.

Clinical evidence (example)

Source	Indication	Weight (example)
Peer-reviewed literature on linear endocutters	Bariatric, colorectal, thoracic	Pivotal
Predicate and equivalence data (ECHELON FLEX, GST reloads)	All	Supportive
PMS data, complaints, vigilance	All	Supportive
PMCF study	Colorectal	Planned

Clinical investigation

Staples are a well-established technology listed in Art. 52(4); a clinical investigation may be waived under Art. 61(6)(b) when the evaluation rests on sufficient clinical data. Example conclusion; not a real CER.

Benefit–risk (example)

Benefit–risk acceptable for the intended use; staple-line leak and bleeding are the main risks and are addressed by labelling and reload selection.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN

SPECIMEN