SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Manufacturer (example) (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

EU DECLARATION OF CONFORMITY

EU declaration of conformity: ECHELON 3000 handle

EU declaration of conformity (MDR Annex IV), also referring to RoHS 2011/65/EU and the Batteries Regulation for the pack · Regulation (EU) 2017/745, Art. 19, Annex IV

Declaration reference	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECHELON 3000 60 mm Standard stapler, ECH60S
Key data	Class not public (research estimate IIa, rule 6)
Scope	Example declaration. The handle's Basic UDI-DI, class and certificate are not public; the specimen leaves them open.

Annex IV elements

Manufacturer and SRN	Legal manufacturer of the ECHELON 3000 system (example holder); SRN not reproduced
Basic UDI-DI	Not public — left open
UDI-DI (information)	10705036035015 (AccessGUDID)
Risk class and rule	To be declared — not public
Union law	Regulation (EU) 2017/745; Directive 2011/65/EU (RoHS); Regulation (EU) 2023/1542 for the battery pack (example)
Standards	EN 60601-1:2006+A13:2024 (harmonised from 17 Jun 2026), IEC 60601-1-2, IEC 62304, IEC 62366-1, EN ISO 14971 (example list)
Notified body and certificate	Notified body (example class) · not issued
Signatory	Example only, not signed

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN

SPECIMEN