SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

### Test laboratory (example issuer class)

Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

TEST REPORT

# EMC test report: ECH60S handle

Electromagnetic disturbances, IEC 60601-1-2:2014+AMD1:2020, professional healthcare facility environment · IEC 60601-1-2:2014+AMD1:2020

| | |
|---|---|
| **Report number** | NOT ISSUED · EXAMPLE |
| **Holder (example)** | Legal manufacturer of the ECHELON 3000 system (example holder) |
| **Product** | ECH60S handle with battery pack |
| **Key data** | CISPR 11 Group 1 Class A · professional healthcare environment (public: IFU) |
| **Scope** | Example scope statement. Emissions and immunity of the battery-powered handle in the operating room, including proximity to RF wireless equipment. |

## Tests (example results)

| Test | Standard | Level | Result (example) |
|---|---|---|---|
| Radiated emissions | CISPR 11 | Group 1 Class A (IFU) | Pass |
| ESD | IEC 61000-4-2 | ±8 kV contact, ±15 kV air | Pass |
| Radiated RF immunity | IEC 61000-4-3 | 3 V/m, 80 MHz–2.7 GHz | Pass |
| Proximity fields from RF wireless | IEC 61000-4-3 / table 9 | 380 MHz–5.8 GHz (IFU) | Pass |
| Power-frequency magnetic field | IEC 61000-4-8 | 30 A/m | Pass |
| Proximity magnetic fields | IEC 61000-4-39 | Table 11 | Pass |

## Essential performance during immunity tests (example)

Essential performance per the IFU: no knife movement without activation of the firing trigger. Test outcome shown is an example.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN