SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Notified body (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

EU TD CERTIFICATE

EU technical documentation assessment certificate: GST60 reloads

EU technical documentation assessment certificate (Annex IX Chapter II), class IIb implantable well-established technology · Regulation (EU) 2017/745, Annex IX, Chapter II; Art. 52(4)

Certificate number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECHELON ENDOPATH reloads with Gripping Surface Technology, 60 mm: GST60W, GST60B, GST60D, GST60G, GST60T
Key data	Class IIb · implantable · staples (WET)
Scope	Example scope statement. Technical documentation of the generic device group 'staple reloads for linear endocutters' assessed on a representative device (GST60B). Covers devices under Basic UDI-DI 0705036a000556X; GST60W to be confirmed.

Devices covered

Reference	UDI-DI (public)	Colour	Closed height
GST60W	10705036014683	White	1.0 mm
GST60B	10705036014706	Blue	1.5 mm
GST60D	10705036014690	Gold	1.8 mm
GST60G	10705036014713	Green	2.0 mm
GST60T	10705036014720	Black	2.3 mm

Basis of the assessment (Annex XII Ch. II item 10)

- Representative device GST60B (example sampling decision under Art. 52(4) second subparagraph)
- Biological evaluation report per EN ISO 10993-1:2025 (specimen in this demo)
- Radiation sterilisation validation per EN ISO 11137-1/-2 (specimen)
- Implant design verification per EN ISO 14630:2024 (example)
- Clinical evaluation report; clinical investigation not required under Art. 61(6)(b) (example conclusion)
- CMR justification for cobalt under Annex I 10.4.2 (example)

Conditions (example)

- SSCP to be validated and uploaded to EUDAMED (Art. 32)
- PSUR at least annually (Art. 86)
- Previous certificate: none (this is a specimen; Annex XII Ch. I 8 requires a reference when a certificate replaces another)

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN