SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Manufacturer (example) (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

PSUR

Periodic safety update report: GST60 reloads

Periodic safety update report, class IIb implantable (at least annually; submitted through EUDAMED to the notified body) · MDR Art. 86, MDCG 2022-21

PSUR reference	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECHELON ENDOPATH GST60 reloads (Basic UDI-DI 0705036a000556X)
Key data	Reporting period: 12 months (example)
Scope	Example scope statement. PMS data for all GST60 reloads sold in the EU, EEA and Northern Ireland.

Summary (example data)

Units sold in period	Not shown (commercial; example placeholder)
Serious incidents reported	Example: 4
FSCAs	Example: 1 (lot-scoped; see change event E1)
Trend report (Art. 88)	Not triggered (example)
PMCF	Colorectal study planned (example)

Complaints by IMDRF device problem (example)

Device problem (IMDRF Annex A, term only)	Count (example)	Rate trend
Failure to form staple	7	Stable
Misfire	5	Stable
Difficult to load	3	Down
Packaging problem	2	Stable

Conclusion (example)

Benefit–risk unchanged; no new risks identified. Example only.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN