SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Manufacturer (example) (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

RISK MANAGEMENT REPORT

Risk management report: ECH60S stapling system

Risk management report per EN ISO 14971:2019+A11:2021 (review before release) · EN ISO 14971:2019+A11:2021

Report number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECH60S handle with GST60 reloads and battery pack
Key data	Overall residual risk: acceptable (example)
Scope	Example scope statement. Covers handle, battery, software, reloads and staples as a system, including combination with staple-line reinforcement.

Top hazards (example)

Hazard	Cause	Control	Evidence	Residual (example)
Staple-line leak	Wrong reload for tissue, malformation	Reload chart, lock-out, bench tests	Staple formation, burst pressure	Acceptable
Bleeding	Inadequate haemostasis on vessels	White reload, IFU warnings	Preclinical haemostasis	Acceptable
Device stops mid-firing	Battery or software fault	Manual override, fault handling	IEC 60601-1, IEC 62304	Acceptable
Infection	Loss of sterility	Validated sterilisation, packaging	ISO 11137, ISO 11607	Acceptable
MR heating	Staples in MR scanner	MR Conditional labelling	ASTM F2503 assessment	Acceptable
Battery hazard at disposal	Undrained lithium pack	Leave installed to drain (IFU)	IEC 60086-4	Acceptable

Review statement (example)

The risk management plan was implemented, overall residual risk is acceptable in relation to the benefit, and methods are in place to collect production and post-production information. Example statement.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN