SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Sterilisation validation laboratory (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

VALIDATION REPORT

Radiation sterilisation validation report: GST60 reloads

Sterilisation validation by gamma irradiation, dose substantiation (VDmax25) per EN ISO 11137-2 with process qualification per EN ISO 11137-1 · EN ISO 11137-2:2015+A1:2023

Report number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	GST60 reloads in final sterile barrier packaging
Key data	Sterilisation dose 25 kGy minimum · SAL 10^{-6} (public: K183435); all test values below are examples
Scope	Example scope statement. Validation of the gamma sterilisation process for the product family 'staple reloads 60 mm'. Routine dose audits quarterly.

Method

Approach	VDmax25 dose substantiation (EN ISO 11137-2 clause 9) — example
Product family	60 mm GST reloads, adopted on the most challenging configuration (example: GST60T)
Bioburden method	EN ISO 11737-1, recovery efficiency determined (example)
Sterility test	EN ISO 11737-2 (example)

Results (example values)

Step	Criterion	Result (example)
Average bioburden, 3 lots × 10 units	≤ 1000 CFU	38 CFU
Verification dose	Per table 9	VDmax25 = 6.4 kGy (example)
Sterility tests at verification dose	≤ 1 positive of 10	0 of 10
Dose mapping, min / max	≥ 25 kGy / ≤ max acceptable	26.1 / 38.4 kGy (example)
Material qualification at max dose	Functional after ageing	Pass (example)

Conclusion (example)

A minimum dose of 25 kGy achieves a sterility assurance level of 10^{-6} for the product family. Quarterly dose audits are required. Example conclusion; not a real validation.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN