SPECIMEN · STRUCTURE ONLY · NOT ISSUED. Example data in a Spherity concept demonstration. No issuing body was involved; no real certificate number, notified-body number or mark is shown. Not provided, reviewed or endorsed by Johnson & Johnson MedTech / Ethicon.

Manufacturer (example) (example issuer class)
Name, address and identification number of the issuer intentionally not shown

NO MARK
NO NUMBER

USABILITY REPORT

Usability engineering summary: ECHELON 3000 with GST60 reloads

Usability engineering file summary and summative evaluation report (IEC 62366-1:2015+AMD1:2020) · IEC 62366-1:2015+AMD1:2020

Report number	NOT ISSUED · EXAMPLE
Holder (example)	Legal manufacturer of the ECHELON 3000 system (example holder)
Product	ECH60S handle, GST60 reloads, IFU and reload chart
Key data	Users: surgeons, surgical residents, scrub nurses (example)
Scope	Example scope statement. Use-related risks of loading, reload selection, articulation, firing, knife return and battery installation.

Hazard-related use scenarios (example)

Scenario	Potential harm	Mitigation	Summative result (example)
Wrong reload colour for tissue	Malformed staples, leak	Colour coding, IFU chart, package label	15/15 correct
45 mm reload in 60 mm handle	Misfire	Mechanical incompatibility, lock-out	Not possible
Firing with used reload	No staples placed	Lock-out	15/15 detected
Battery not installed	Delay	Haptic and audible feedback	15/15 recovered
Knife not returning	Tissue damage	Manual override	14/15 first attempt (example)

Conclusion (example)

Residual use-related risk acceptable; one IFU wording change on the manual override. Example only.

Signature (example, not signed)

Date: example only

SPHERITY Specimen generated by Spherity for a concept demonstration on 2026-10-16. Public facts are cited in the viewer; all results, numbers, dates and names of parties are examples.

SPECIMEN