

Table of contents

\\FORSKNING

Johanna

TSSfinal

TSS_coil_QA_fMRI_23-07-03_D

<p> localizer_D localizer_D tfl_sag_2p5mmISO_largeFOV tfl_sag_2p5mmISO_largeFOV_refv550 DREAM2D_sag_2p5mmISO_largeFOV DREAM2D_sag_2p5mmISO_mediumFOV DREAM2D_sag_2p5mmISO_mediumFOV_refv367 coilQA_sag_FH_1p5x1p5x2mm_largeFOV coilQA_sag_FH_1p5x1p5x2mm_smallFOV coilQA_tra_RL_0p5x0p5x5mm gre_2mmISO_multichannel_uncomb localizer_D t1_mp2rage_cor_nonSelHS1_0.7iso tfl_sag_2p5mmISO_largeFOV_refv550 ep2d_lgFOVc45_0p85_R2PF6_TE30_BW1132 dzne_ep3d_fmri_EPI16 dzne_ep3d_fmri_EPI10 gre_B0_sag gre_B0_tra </p>
--

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmRI_23-07-03_D\localizer_D

TA: 0:18 PM: REF Voxel size: 0.5×0.5×5.0 mmPAT: Off Rel. SNR: 1.00 : qfl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	3
Dist. factor	50 %
Position	R2.9 A15.4 F49.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	3
Dist. factor	50 %
Position	R15.8 A13.6 F49.8 mm
Orientation	Coronal
Phase enc. dir.	F >> H
Slice group	3
Slices	3
Dist. factor	50 %
Position	R11.0 A7.5 F38.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	25 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	7.7 ms
TE	3.67 ms
Averages	1
Concatenations	9
Filter	Distortion Corr.(3D), Elliptical filter
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	7.7 ms
TE	3.67 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement
-----------------	------------------

Resolution - Common

FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	3
Dist. factor	50 %
Position	R2.9 A15.4 F49.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	3
Dist. factor	50 %
Position	R15.8 A13.6 F49.8 mm
Orientation	Coronal
Phase enc. dir.	F >> H
Slice group	3
Slices	3
Dist. factor	50 %
Position	R11.0 A7.5 F38.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	7.7 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	9

Geometry - AutoAlign

Slice group	1
Position	R2.9 A15.4 F49.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2

Geometry - AutoAlign

Position	R15.8 A13.6 F49.8 mm
Orientation	Coronal
Phase enc. dir.	F >> H
Slice group	3
Position	R11.0 A7.5 F38.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R2.9 A15.4 F49.8
R	2.9 mm
A	15.4 mm
F	49.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim CT

Tim CT mode	Off
Slices	3
Slice thickness	5.0 mm
Dist. factor	50 %
FoV read	280 mm
FoV phase	100.0 %
Segments	1

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	H >> F
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.0 A25.1 F32.6 mm
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	175 mm
! R >> L	201 mm
! F >> H	259 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1

System - Tx/Rx

Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	400.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7.7 ms
Concatenations	9
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	280 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	9

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	Active

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\localizer_D

TA: 7.0 s PM: REF Voxel size: 0.5×0.5×5.0 mmPAT: Off Rel. SNR: 1.00 : qfl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	3
Dist. factor	50 %
Position	R15.8 A44.6 F50.5 mm
Orientation	Coronal
Phase enc. dir.	F >> H
AutoAlign	---
Phase oversampling	25 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	7.7 ms
TE	3.67 ms
Averages	1
Concatenations	3
Filter	Distortion Corr.(3D), Elliptical filter
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Contrast - Common

TR	7.7 ms
TE	3.67 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	3
Dist. factor	50 %
Position	R15.8 A44.6 F50.5 mm
Orientation	Coronal
Phase enc. dir.	F >> H
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	7.7 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	3

Geometry - AutoAlign

Slice group	1
Position	R15.8 A44.6 F50.5 mm
Orientation	Coronal
Phase enc. dir.	F >> H
AutoAlign	---
Initial Position	R15.8 A44.6 F50.5
R	15.8 mm
A	44.6 mm
F	50.5 mm
Initial Rotation	89.80 deg
Initial Orientation	Coronal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim CT

Tim CT mode	Off
Slices	3
Slice thickness	5.0 mm
Dist. factor	50 %
FoV read	280 mm
FoV phase	100.0 %
Segments	1

System - Miscellaneous

Positioning mode	REF
------------------	-----

System - Miscellaneous

Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	H >> F
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.0 A25.1 F32.6 mm
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	175 mm
! R >> L	201 mm
! F >> H	259 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	400.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7.7 ms
Concatenations	3
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	280 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	3

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	Active
RF pulse type	Normal
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\tfl_sag_2p5mmISO_largeFOV

TA: 1:10 PM: ISO Voxel size: 2.4×2.4×2.5 mmPAT: Off Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	28
Dist. factor	100 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	28
Dist. factor	100 %
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
TR	34420.0 ms
TE	2.31 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Contrast - Common

TR	34420.0 ms
TE	2.31 ms
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
Base resolution	144
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	28
Dist. factor	100 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	28
Dist. factor	100 %
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
TR	34420.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 F2.9
L	2.4 mm
A	45.3 mm
F	2.9 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	ISO
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P

System - Miscellaneous

Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

Sequence - Part 2

RF spoiling	On
Turbo factor	88

Sequence - Assistant

Mode	Off
------	-----

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L1.2 A45.3 F2.9 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	208 mm
F >> H	341 mm
R >> L	140 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
? Ref. amplitude 1H	0.000 V

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	Off
Dimension	2D
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Interleaved
Echo spacing	5.1 ms
Bandwidth	400 Hz/Px

Sequence - Part 2

RF pulse type	Low SAR
Gradient mode	Normal
Excitation	Slice-sel.

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmri_23-07-03_D\tfl_sag_2p5mmISO_largeFOV_refv
550

TA: 1:10 PM: ISO Voxel size: 2.4×2.4×2.5 mmPAT: Off Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	28
Dist. factor	100 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	28
Dist. factor	100 %
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
TR	34420.0 ms
TE	2.31 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	34420.0 ms
TE	2.31 ms
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
Base resolution	144
Phase resolution	100 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	28
Dist. factor	100 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	28
Dist. factor	100 %
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
TR	34420.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 F2.9
L	2.4 mm
A	45.3 mm
F	2.9 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	ISO
------------------	-----

System - Miscellaneous

Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L1.2 A45.3 F2.9 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	208 mm
F >> H	341 mm
R >> L	140 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	Off
Dimension	2D
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Interleaved
Echo spacing	5.1 ms
Bandwidth	400 Hz/Px

Sequence - Part 2

RF pulse type	Low SAR
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	88

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmri_23-07-03_D\DREAM2D_sag_2p5mmISO_largeFOV

TA: 1:08 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: Off Rel. SNR: 1.00 : 727f999

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	11
Dist. factor	0 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
FoV read	384 mm
FoV phase	47.4 %
Slice thickness	2.5 mm
TR	6000 ms
TE 1	2.04 ms
TE 2	3.18 ms
Averages	1
Concatenations	11
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	6000 ms
TE 1	2.04 ms
TE 2	3.18 ms
Flip angle 1	50 deg
Flip angle 2	6 deg

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	384 mm
FoV phase	47.4 %
Slice thickness	2.5 mm
Base resolution	152
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
--------------	-----

Resolution - Filter Image

Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	11
Dist. factor	0 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	384 mm
FoV phase	47.4 %
Slice thickness	2.5 mm
TR	6000 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	11

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off

System - Adjustments

Adjustment Tolerance	Auto
----------------------	------

System - Adjust Volume

! Position	L0.9 A43.0 F14.6 mm
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	166 mm
! R >> L	41 mm
! F >> H	350 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Multi-slice mode	Sequential
Bandwidth	910 Hz/Px

Sequence - Part 2

Echo train duration	357 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	72

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Special

Preparation scans	2
Preparation loops	0
Sample T1	2000 ms
RF-Duration	800 us
Prep RF-Duration	400 us
Time-BW-Product	3.2
Timing Scheme	STE*
Mixing Time	1140 us
FFT Scale	10
Calculate FlipMap	On
HDR DICOMs	Off
Scale risetime	1.20

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\IDREAM2D_sag_2p5mmISO_mediu
mFOV

TA: 1:08 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: Off Rel. SNR: 1.00 : 727f999

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	11
Dist. factor	0 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
TR	6000 ms
TE 1	2.04 ms
TE 2	3.10 ms
Averages	1
Concatenations	11
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	6000 ms
TE 1	2.04 ms
TE 2	3.10 ms
Flip angle 1	50 deg
Flip angle 2	6 deg

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
Base resolution	80
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
--------------	-----

Resolution - Filter Image

Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	11
Dist. factor	0 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
TR	6000 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	11

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off

System - Adjustments

Adjustment Tolerance	Auto
----------------------	------

System - Adjust Volume

! Position	L0.9 A45.6 F3.9 mm
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	166 mm
! R >> L	41 mm
! F >> H	199 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Multi-slice mode	Sequential
Bandwidth	1010 Hz/Px

Sequence - Part 2

Echo train duration	386 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	80

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Special

Preparation scans	2
Preparation loops	0
Sample T1	2000 ms
RF-Duration	800 us
Prep RF-Duration	400 us
Time-BW-Product	3.2
Timing Scheme	STE*
Mixing Time	1060 us
FFT Scale	10
Calculate FlipMap	On
HDR DICOMs	Off
Scale risetime	1.20

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\IDREAM2D_sag_2p5mmISO_mediu
mFOV_refv367

TA: 1:08 PM: FIX Voxel size: 2.5×2.5×2.5 mmPAT: Off Rel. SNR: 1.00 : 727f999

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	11
Dist. factor	0 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
TR	6000 ms
TE 1	2.04 ms
TE 2	3.10 ms
Averages	1
Concatenations	11
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	6000 ms
TE 1	2.04 ms
TE 2	3.10 ms
Flip angle 1	50 deg
Flip angle 2	6 deg

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1

Resolution - Common

FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
Base resolution	80
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
--------------	-----

Resolution - Filter Image

Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	11
Dist. factor	0 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.5 mm
TR	6000 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	11

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off

System - Adjustments

Adjustment Tolerance	Auto
----------------------	------

System - Adjust Volume

! Position	L0.9 A45.6 F3.9 mm
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	166 mm
! R >> L	41 mm
! F >> H	199 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	367.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Multi-slice mode	Sequential
Bandwidth	1010 Hz/Px

Sequence - Part 2

Echo train duration	386 ms
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	80

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Special

Preparation scans	2
Preparation loops	0
Sample T1	2000 ms
RF-Duration	800 us
Prep RF-Duration	400 us
Time-BW-Product	3.2
Timing Scheme	STE*
Mixing Time	1060 us
FFT Scale	10
Calculate FlipMap	On
HDR DICOMs	Off
Scale risetime	1.20

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmRI_23-07-03_D\coilQA_sag_FH_1p5x1p5x2mm_largeFOV

TA: 3:20 PM: FIX Voxel size: 1.5×1.5×2.0 mmRel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	13
Dist. factor	20 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	---
Phase oversampling	0 %
FoV read	384 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	30.0 ms
TE	6.0 ms
Averages	2
Concatenations	13
Filter	None
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	30.0 ms
TE	6.0 ms
TD	0 ms
MTC	Off
Flip angle	12 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	2
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	384 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	256
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
--------------	-----

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	13
Dist. factor	20 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
FoV read	384 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	30.0 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	13

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	90.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L2.0 A36.5 F42.5 mm
! Orientation	Sagittal
! Rotation	77.10 deg
! F >> H	269 mm
! A >> P	68 mm
! R >> L	32 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Physio - Signal1

1st Signal/Mode	None
TR	30.0 ms
Concatenations	13

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	200 Hz/Px

Sequence - Part 2

Gradient mode	Fast
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Special

ICE program	IceProgramCoilUtils
Prep. scans duration	0 ms
Optimal SNR	On
GFactor	On
Rx coil diode switching	On

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\coilQA_sag_FH_1p5x1p5x2mm_smallFOV

TA: 1:40 PM: FIX Voxel size: 1.5×1.5×2.0 mmRel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	13
Dist. factor	20 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	30.0 ms
TE	6.0 ms
Averages	2
Concatenations	13
Filter	None
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	30.0 ms
TE	6.0 ms
TD	0 ms
MTC	Off
Flip angle	12 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	2
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	128
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
--------------	-----

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	13
Dist. factor	20 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	30.0 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	13

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	H >> F
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	90.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm

System - Adjustments

Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L2.0 A36.5 F42.5 mm
! Orientation	Sagittal
! Rotation	77.10 deg
! F >> H	269 mm
! A >> P	68 mm
! R >> L	32 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Physio - Signal1

1st Signal/Mode	None
TR	30.0 ms
Concatenations	13

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	200 Hz/Px

Sequence - Part 2

Gradient mode	Fast
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Special

ICE program	IceProgramCoilUtils
Prep. scans duration	0 ms
Optimal SNR	On
GFactor	On
Rx coil diode switching	On

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\coilQA_tra_RL_0p5x0p5x5mm

TA: 2:41 PM: FIX Voxel size: 0.5×0.5×5.0 mmRel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	7
Dist. factor	300 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	30.0 ms
TE	6.0 ms
Averages	2
Concatenations	7
Filter	None
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	30.0 ms
TE	6.0 ms
TD	0 ms
MTC	Off
Flip angle	12 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	2
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	192 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	384
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off

Resolution - Filter Image

Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	7
Dist. factor	300 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Transversal
Phase enc. dir.	R >> L
FoV read	192 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	30.0 ms
Multi-slice mode	Sequential
Series	Ascending
Concatenations	7

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Transversal
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	90.00 deg
Initial Orientation	Transversal

Geometry - Saturation

Fat suppr.	None
Water suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Adaptive Combine
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off

System - Adjustments

Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L2.0 A36.5 F42.5 mm
! Orientation	Sagittal
! Rotation	77.10 deg
! F >> H	269 mm
! A >> P	68 mm
! R >> L	32 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Physio - Signal1

1st Signal/Mode	None
TR	30.0 ms
Concatenations	7

Sequence - Part 1

Introduction	Off
Dimension	2D
Contrasts	1
Multi-slice mode	Sequential
Bandwidth	200 Hz/Px

Sequence - Part 2

Gradient mode	Fast
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Special

ICE program	IceProgramCoilUtils
Prep. scans duration	0 ms
Optimal SNR	On
GFactor	On
Rx coil diode switching	On

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmRI_23-07-03_D\gre_2mmISO_multichannel_uncom b
TA: 1:21 PM: FIX Voxel size: 2.0×2.0×2.0 mmPAT: Off Rel. SNR: 1.00 : fl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	44
Dist. factor	0 %
Position	L2.4 A41.7 F36.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	384 mm
FoV phase	75.0 %
Slice thickness	2.0 mm
TR	550.0 ms
TE	3.80 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Contrast - Common

TR	550.0 ms
TE	3.80 ms
MTC	Off
Magn. preparation	None
Flip angle	30 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	384 mm
FoV phase	75.0 %
Slice thickness	2.0 mm
Base resolution	192
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	44
Dist. factor	0 %
Position	L2.4 A41.7 F36.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	384 mm
FoV phase	75.0 %
Slice thickness	2.0 mm
TR	550.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L2.4 A41.7 F36.0 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A41.7 F33.0
L	2.4 mm
A	41.7 mm
F	33.0 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim CT

Tim CT mode	Off
Slices	44
Slice thickness	2.0 mm
Dist. factor	0 %
FoV read	384 mm
FoV phase	75.0 %
Segments	1

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	On
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L2.4 A41.7 F36.0 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	288 mm
F >> H	384 mm
R >> L	88 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Physio - Signal1

1st Signal/Mode	None
TR	550.0 ms
Concatenations	1
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	384 mm
FoV phase	75.0 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	On
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	None
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slice-sel.
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmRI_23-07-03_D\localizer_D

TA: 0:18 PM: REF Voxel size: 0.5×0.5×5.0 mmPAT: Off Rel. SNR: 1.00 : qfl

Properties

Prio recon	On
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	On
Load images to graphic segments	On
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	3
Dist. factor	50 %
Position	R2.9 A15.4 F49.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	3
Dist. factor	50 %
Position	R15.8 A45.8 F48.9 mm
Orientation	Coronal
Phase enc. dir.	F >> H
Slice group	3
Slices	3
Dist. factor	50 %
Position	R11.0 A7.5 F38.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	25 %
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	7.7 ms
TE	3.67 ms
Averages	1
Concatenations	9
Filter	Distortion Corr.(3D), Elliptical filter
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	7.7 ms
TE	3.67 ms
TD	0 ms
MTC	Off
Magn. preparation	None
Flip angle	20 deg
Fat suppr.	None
Water suppr.	None
SWI	Off

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magnitude
Measurements	1

Contrast - Dynamic

Multiple series	Each measurement
-----------------	------------------

Resolution - Common

FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
Base resolution	256
Phase resolution	75 %
Phase partial Fourier	Off
Interpolation	On

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	On

Geometry - Common

Slice group	1
Slices	3
Dist. factor	50 %
Position	R2.9 A15.4 F49.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	3
Dist. factor	50 %
Position	R15.8 A45.8 F48.9 mm
Orientation	Coronal
Phase enc. dir.	F >> H
Slice group	3
Slices	3
Dist. factor	50 %
Position	R11.0 A7.5 F38.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
FoV read	280 mm
FoV phase	100.0 %
Slice thickness	5.0 mm
TR	7.7 ms
Multi-slice mode	Sequential
Series	Interleaved
Concatenations	9

Geometry - AutoAlign

Slice group	1
Position	R2.9 A15.4 F49.8 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2

Geometry - AutoAlign

Position	R15.8 A45.8 F48.9 mm
Orientation	Coronal
Phase enc. dir.	F >> H
Slice group	3
Position	R11.0 A7.5 F38.4 mm
Orientation	Transversal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R2.9 A15.4 F49.8
R	2.9 mm
A	15.4 mm
F	49.8 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None
Water suppr.	None
Special sat.	None

Geometry - Tim CT

Tim CT mode	Off
Slices	3
Slice thickness	5.0 mm
Dist. factor	50 %
FoV read	280 mm
FoV phase	100.0 %
Segments	1

System - Miscellaneous

Positioning mode	REF
Table position	F
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	H >> F
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Off - AutoCoilSelect

System - Adjustments

B0 Shim mode	Tune up
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.0 A25.1 F32.6 mm
! Orientation	Transversal
! Rotation	0.00 deg
! A >> P	175 mm
! R >> L	201 mm
! F >> H	259 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1

System - Tx/Rx

Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	400.000 V

Physio - Signal1

1st Signal/Mode	None
TR	7.7 ms
Concatenations	9
Segments	1

Physio - Cardiac

Magn. preparation	None
Fat suppr.	None
Dark blood	Off
FoV read	280 mm
FoV phase	100.0 %
Phase resolution	75 %

Physio - PACE

Resp. control	Off
Concatenations	9

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Liver registration	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Soft Tissue

Wash - In	Off
Wash - Out	Off
TTP	Off
PEI	Off
MIP - time	Off
Measurements	1

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	On
Dimension	2D
Phase stabilisation	Off
Asymmetric echo	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Sequential
Bandwidth	320 Hz/Px

Sequence - Part 2

Segments	1
Acoustic noise reduction	Active

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D1t1_mp2rage_cor_nonSelHS1_0.7is o
TA: 8:47 PM: FIX Voxel size: 0.7×0.7×0.7 mmPAT: 2 Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R1.7 A47.5 F5.9 mm
Orientation	C > T-12.1 > S-0.1
Phase enc. dir.	R >> L
AutoAlign	---
Phase oversampling	0 %
Slice oversampling	0.0 %
Slices per slab	192
FoV read	260 mm
FoV phase	65.8 %
Slice thickness	0.70 mm
TR	5000.0 ms
TE	2.15 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	5000.0 ms
TE	2.15 ms
Magn. preparation	Non-sel. IR
TI 1	700 ms
TI 2	2400 ms
Flip angle 1	4.0 deg
Flip angle 2	5.0 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	260 mm
FoV phase	65.8 %
Slice thickness	0.70 mm
Base resolution	368
Phase resolution	100 %
Slice resolution	100 %

Resolution - Common

Phase partial Fourier	6/8
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	28
Accel. factor 3D	1
Reference scan mode	Integrated

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Dist. factor	50 %
Position	R1.7 A47.5 F5.9 mm
Orientation	C > T-12.1 > S-0.1
Phase enc. dir.	R >> L
Slice oversampling	0.0 %
Slices per slab	192
FoV read	260 mm
FoV phase	65.8 %
Slice thickness	0.70 mm
TR	5000.0 ms
Multi-slice mode	Single shot
Series	Ascending
Concatenations	1

Geometry - AutoAlign

Slab group	1
Position	R1.7 A47.5 F5.9 mm
Orientation	C > T-12.1 > S-0.1
Phase enc. dir.	R >> L
AutoAlign	---
Initial Position	R1.7 A47.5 F2.9
R	1.7 mm
A	47.5 mm
F	2.9 mm
Initial Rotation	-0.70 deg
Initial Orientation	C > T
C > T	-12.1
> S	-0.1

Geometry - Navigator

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L1.0 A51.0 H10.9 mm
! Orientation	T > C10.5 > S2.3
! Rotation	-0.20 deg
! A >> P	51 mm
! R >> L	59 mm
! F >> H	145 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Physio - Signal1

1st Signal/Mode	None
TR	5000.0 ms
Concatenations	1

Physio - Cardiac

Magn. preparation	Non-sel. IR
TI 1	700 ms
TI 2	2400 ms
Fat suppr.	None
Dark blood	Off
FoV read	260 mm
FoV phase	65.8 %
Phase resolution	100 %

Physio - PACE

Resp. control	Off
Concatenations	1

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	Off

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Single shot
Echo spacing	6.4 ms
Bandwidth	220 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Normal*
Excitation	Non-sel.
RF spoiling	On
Incr. Gradient spoiling	Off
Turbo factor	192

Sequence - Nuclei

TX/RX Nucleus	1H
TX/RX delta frequency	0 Hz
TX Nucleus	None
TX delta frequency	0 Hz
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H

Sequence - Special

Use Custom Inversion	On
Inv pulse type	HS1
HS pulse dur	10240 us
HS pulse offset	0 Hz
HS flip angle	360 deg
TR_FOCI B1	0.00 uT
Echo Spacing	6400 us
Denoise Weighting	100

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmRI_23-07-03_D\tfl_sag_2p5mmISO_largeFOV_refv
550

TA: 1:10 PM: ISO Voxel size: 2.4×2.4×2.5 mmPAT: Off Rel. SNR: 1.00 : tfl

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	28
Dist. factor	100 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	28
Dist. factor	100 %
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
TR	34420.0 ms
TE	2.31 ms
Averages	1
Concatenations	1
Filter	Distortion Corr.(3D)
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	34420.0 ms
TE	2.31 ms
Magn. preparation	None
Flip angle	10 deg
Fat suppr.	None
Water suppr.	None

Contrast - Dynamic

Averages	1
Reconstruction	Magnitude
Measurements	1
Multiple series	Each measurement

Resolution - Common

FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
Base resolution	144
Phase resolution	100 %
Phase partial Fourier	Off

Resolution - Common

Interpolation	Off
---------------	-----

Resolution - iPAT

PAT mode	None
----------	------

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	On
Mode	3D
Unfiltered images	On
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	28
Dist. factor	100 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Slices	28
Dist. factor	100 %
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	340 mm
FoV phase	61.1 %
Slice thickness	2.5 mm
TR	34420.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
Slice group	2
Position	R0.1 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 F2.9
L	2.4 mm
A	45.3 mm
F	2.9 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

System - Miscellaneous

Positioning mode	ISO
------------------	-----

System - Miscellaneous

Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

Position	L1.2 A45.3 F2.9 mm
Orientation	Sagittal
Rotation	0.00 deg
A >> P	208 mm
F >> H	341 mm
R >> L	140 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Inline - Common

Subtract	Off
Measurements	1
StdDev	Off
Save original images	On

Inline - MIP

MIP-Sag	Off
MIP-Cor	Off
MIP-Tra	Off
MIP-Time	Off
Save original images	On

Inline - Composing

Distortion Corr.	On
Mode	3D
Unfiltered images	On

Sequence - Part 1

Introduction	Off
Dimension	2D
Asymmetric echo	Allowed
Flow comp.	No
Multi-slice mode	Interleaved
Echo spacing	5.1 ms
Bandwidth	400 Hz/Px

Sequence - Part 2

RF pulse type	Low SAR
Gradient mode	Normal
Excitation	Slice-sel.
RF spoiling	On
Turbo factor	88

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_Dlep2d_IgFOVc45_0p85_R2PF6_TE3
0_BW1132

TA: 10:10 PM: FIX Voxel size: 0.9×0.9×3.0 mmPAT: 2 Rel. SNR: 1.00 : epfid

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	16
Dist. factor	0 %
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	1290 ms
TE	30.00 ms
Multi-band accel. factor	1
Filter	Raw filter
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;F21-24

Contrast - Common

TR	1290 ms
TE	30.00 ms
MTC	Off
Magn. preparation	None
Flip angle	110 deg
Fat suppr.	None

Contrast - Dynamic

Averaging mode	Long term
Reconstruction	Magnitude
Measurements	466
Delay in TR	0 ms
Multiple series	Off

Resolution - Common

FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
Base resolution	192
Phase resolution	100 %
Phase partial Fourier	6/8
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Accel. factor PE	2
Ref. lines PE	52

Resolution - iPAT

Reference scan mode	GRE
---------------------	-----

Resolution - Filter Image

Distortion Corr.	Off
Prescan Normalize	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off
Hamming	Off

Geometry - Common

Slice group	1
Slices	16
Dist. factor	0 %
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	1290 ms
Multi-slice mode	Interleaved
Series	Ascending
Multi-band accel. factor	1

Geometry - AutoAlign

Slice group	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R2.9 A53.3 F9.2
R	2.9 mm
A	53.3 mm
F	9.2 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	12.3
> S	0.0

Geometry - Saturation

Fat suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.6 A46.9 F8.0 mm
! Orientation	T > C12.3
! Rotation	0.00 deg
! A >> P	40 mm
! R >> L	52 mm
! F >> H	60 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Physio - Signal1

1st Signal/Mode	None
TR	1290 ms
Multi-band accel. factor	1

BOLD

GLM Statistics	Off
Dynamic t-maps	Off
Ignore meas. at start	0
Ignore after transition	0
Model transition states	On
Temp. highpass filter	On
Threshold	4.00
Paradigm size	20
Meas[1]	Baseline
Meas[2]	Baseline
Meas[3]	Baseline
Meas[4]	Baseline
Meas[5]	Baseline
Meas[6]	Baseline
Meas[7]	Baseline
Meas[8]	Baseline
Meas[9]	Baseline
Meas[10]	Baseline
Meas[11]	Active
Meas[12]	Active
Meas[13]	Active
Meas[14]	Active
Meas[15]	Active
Meas[16]	Active
Meas[17]	Active
Meas[18]	Active
Meas[19]	Active
Meas[20]	Active
Motion correction	Off
Spatial filter	Off
Measurements	466
Delay in TR	0 ms
Multiple series	Off

Sequence - Part 1

Introduction	Off
Contrasts	1
Flow comp.	No
Multi-slice mode	Interleaved
Free echo spacing	Off
Echo spacing	1 ms
Bandwidth	1132 Hz/Px

Sequence - Part 2

EPI factor	192
Gradient mode	Fast
RF spoiling	Off

Sequence - Special

Excite pulse duration	3840 us
SENSE1 coil combine	Off
Invert RO/PE polarity	Off
PF omits higher k-space	Off
Disable freq. update	Off
Force equal slice timing	Off
FFT scale factor	1.00
GRE iPAT ref. FA	12.0 deg
Physio recording	Legacy
Triggering scheme	Standard

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmri_23-07-03_D\dzne_ep3d_fmri_EPI16

TA: 10:14 PM: FIX Voxel size: 0.9×0.9×3.0 mmPAT: 2 Rel. SNR: 1.00 : ep3d f015566

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	16
FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	2300 ms
TE 1	11.20 ms
Averages	1
Multi-echo Shots	1
Filter	Raw filter
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Contrast - Common

TR	2300 ms
TE 1	11.20 ms
Multi-echo spacing	17.18 ms
MTC	Off
Magn. preparation	None
TI	900 ms
Flip angle	17 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	264
Pause after meas.	0.0 s

Resolution - Common

FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Acc. factor PE	2
Ref. lines PE	52
Acc. factor 3D	1
Ref. lines 3D	16
Reference Scan Mode	GRE/separate

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	16
FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	2300 ms
Multi-slice mode	Interleaved
Series	Ascending
Multi-echo Shots	1

Geometry - AutoAlign

Slab group	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R2.9 A53.3 F9.2
R	2.9 mm
A	53.3 mm
F	9.2 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	12.3
> S	0.0

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P

System - Miscellaneous

Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

Sequence - Special

EPI rise time factor	1.10
Mosaic DICOMs	On

Sequence - Assistant

Mode	Off
------	-----

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Freq. adjust	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.6 A46.9 F8.0 mm
! Orientation	T > C12.3
! Rotation	0.00 deg
! A >> P	40 mm
! R >> L	52 mm
! F >> H	60 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	1
Multi-slice mode	Interleaved
Echo spacing	1.01 ms
Bandwidth	1240 Hz/Px

Sequence - Part 2

EPI factor	16
Segmentation	6
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On

Sequence - Special

PATRef FA	5 deg
RF duration	2560 us
RF BWT product	30
Ernst T1	1300 ms
PATRef prep. shots	200
Volume dummy shots	0
Dummy Measurements	0
CHECK FLIP ANGLE!	On
Integrated PC	Off
Invert PE	Off
Water Exc.	-none-
Phase Correction	per Blade

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fmri_23-07-03_D\dzne_ep3d_fmri_EPI10

TA: 10:17 PM: FIX Voxel size: 0.9×0.9×3.0 mmPAT: 2 Rel. SNR: 1.00 : ep3d f015566

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slab group	1
Slabs	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Slab Scale	-10 %
Slices per slab	16
FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	2800 ms
TE 1	8.04 ms
Averages	1
Multi-echo Shots	1
Filter	Raw filter
Coil elements	A01-04;B05-08;C09-12;D13-16;E17-20;H21-24

Contrast - Common

TR	2800 ms
TE 1	8.04 ms
Multi-echo spacing	11.32 ms
MTC	Off
Magn. preparation	None
TI	900 ms
Flip angle	15 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Long term
Reconstruction	Magnitude
Measurements	218
Pause after meas.	0.0 s

Resolution - Common

FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
Base resolution	192
Phase resolution	100 %
Slice resolution	100 %
Phase partial Fourier	Off
Slice partial Fourier	Off
Interpolation	Off

Resolution - iPAT

PAT mode	GRAPPA
Acc. factor PE	2
Ref. lines PE	52
Acc. factor 3D	1
Ref. lines 3D	16
Reference Scan Mode	GRE/separate

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off
Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	On
Elliptical filter	Off

Geometry - Common

Slab group	1
Slabs	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
Slab Scale	-10 %
Slices per slab	16
FoV read	164 mm
FoV phase	100.0 %
Slice thickness	3.00 mm
TR	2800 ms
Multi-slice mode	Interleaved
Series	Ascending
Multi-echo Shots	1

Geometry - AutoAlign

Slab group	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R2.9 A53.3 F9.2
R	2.9 mm
A	53.3 mm
F	9.2 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	12.3
> S	0.0

Geometry - Saturation

Saturation mode	Standard
Fat suppr.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P

System - Miscellaneous

Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

Sequence - Special

Mosaic DICOMs	On
---------------	----

Sequence - Assistant

Mode	Off
------	-----

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Freq. adjust	Off
Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.6 A46.9 F8.0 mm
! Orientation	T > C12.3
! Rotation	0.00 deg
! A >> P	40 mm
! R >> L	52 mm
! F >> H	60 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	Low
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Sequence - Part 1

Introduction	On
Dimension	3D
Elliptical scanning	Off
Reordering	Linear
Asymmetric echo	Off
Contrasts	1
Multi-slice mode	Interleaved
Echo spacing	1.03 ms
Bandwidth	1302 Hz/Px

Sequence - Part 2

EPI factor	10
Segmentation	10
RF pulse type	Normal
Gradient mode	Fast
Excitation	Slab-sel.
RF spoiling	On

Sequence - Special

PATRef FA	5 deg
RF duration	2560 us
RF BWT product	30
Ernst T1	1300 ms
PATRef prep. shots	200
Volume dummy shots	0
Dummy Measurements	0
Integrated PC	Off
Invert PE	Off
Water Exc.	-none-
Phase Correction	per Blade
EPI rise time factor	1.10

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\gre_B0_sag

TA: 0:59 PM: FIX Voxel size: 1.1×1.1×2.0 mmRel. SNR: 1.00 : fm

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	20
Dist. factor	20 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	150.0 ms
TE 1	3.06 ms
TE 2	4.08 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A01-04;B05-08;C09-12;D13-16

Contrast - Common

TR	150.0 ms
TE 1	3.06 ms
TE 2	4.08 ms
MTC	Off
Flip angle	19 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
Base resolution	190
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off

Resolution - Filter Image

Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	20
Dist. factor	20 %
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	2.0 mm
TR	150.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	L2.4 A45.3 F2.9 mm
Orientation	Sagittal
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	L2.4 A45.3 H0.1
L	2.4 mm
A	45.3 mm
F	0.1 mm
Initial Rotation	0.00 deg
Initial Orientation	Sagittal

Geometry - Saturation

Fat suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	F
Table position	3 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off
Assume Dominant Fat	Off
Assume Silicone	Off

System - Adjustments

Adjustment Tolerance	Auto
----------------------	------

System - Adjust Volume

! Position	L0.6 A46.9 F11.0 mm
! Orientation	T > C12.3
! Rotation	0.00 deg
! A >> P	40 mm
! R >> L	52 mm
! F >> H	60 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	797 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----

\\FORSKNING\Johanna\TSSfinal\TSS_coil_QA_fMRI_23-07-03_D\gre_B0_tra

TA: 0:59 PM: FIX Voxel size: 1.1×1.1×3.0 mmRel. SNR: 1.00 : fm

Properties

Prio recon	Off
Load images to viewer	On
Inline movie	Off
Auto store images	On
Load images to stamp segments	Off
Load images to graphic segments	Off
Auto open inline display	Off
Auto close inline display	Off
Start measurement without further preparation	Off
Wait for user to start	Off
Start measurements	Single measurement

Routine

Slice group	1
Slices	16
Dist. factor	0 %
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Phase oversampling	0 %
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	150.0 ms
TE 1	3.06 ms
TE 2	4.08 ms
Averages	1
Concatenations	1
Filter	None
Coil elements	A01-04;B05-08;C09-12;D13-16

Contrast - Common

TR	150.0 ms
TE 1	3.06 ms
TE 2	4.08 ms
MTC	Off
Flip angle	19 deg
Fat suppr.	None

Contrast - Dynamic

Averages	1
Averaging mode	Short term
Reconstruction	Magn./Phase
Measurements	1
Multiple series	Off

Resolution - Common

FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
Base resolution	190
Phase resolution	100 %
Phase partial Fourier	Off
Interpolation	Off

Resolution - Filter Image

Image Filter	Off
Distortion Corr.	Off

Resolution - Filter Image

Prescan Normalize	Off
Normalize	Off
B1 filter	Off

Resolution - Filter Rawdata

Raw filter	Off
Elliptical filter	Off

Geometry - Common

Slice group	1
Slices	16
Dist. factor	0 %
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
FoV read	200 mm
FoV phase	100.0 %
Slice thickness	3.0 mm
TR	150.0 ms
Multi-slice mode	Interleaved
Series	Interleaved
Concatenations	1

Geometry - AutoAlign

Slice group	1
Position	R2.9 A53.3 F9.2 mm
Orientation	T > C12.3
Phase enc. dir.	A >> P
AutoAlign	---
Initial Position	R2.9 A53.3 F9.2
R	2.9 mm
A	53.3 mm
F	9.2 mm
Initial Rotation	0.00 deg
Initial Orientation	T > C
T > C	12.3
> S	0.0

Geometry - Saturation

Fat suppr.	None
Special sat.	None

System - Miscellaneous

Positioning mode	FIX
Table position	H
Table position	0 mm
MSMA	S - C - T
Sagittal	R >> L
Coronal	A >> P
Transversal	F >> H
Coil Combine Mode	Sum of Squares
Save uncombined	Off
Matrix Optimization	Off
AutoAlign	---
Coil Select Mode	Default

System - Adjustments

B0 Shim mode	Brain
B1 Shim mode	TrueForm
Confirm freq. adjustment	Off

System - Adjustments

Assume Dominant Fat	Off
Assume Silicone	Off
Adjustment Tolerance	Auto

System - Adjust Volume

! Position	L0.6 A46.9 F8.0 mm
! Orientation	T > C12.3
! Rotation	0.00 deg
! A >> P	40 mm
! R >> L	52 mm
! F >> H	60 mm
Reset	Off

System - Tx/Rx

Frequency 1H	297.098817 MHz
Correction factor	1
Gain	High
Img. Scale Cor.	1.000
Reset	Off
! Ref. amplitude 1H	550.000 V

Sequence - Part 1

Introduction	On
Dimension	2D
Asymmetric echo	Off
Contrasts	2
Flow comp.	No
Multi-slice mode	Interleaved
Bandwidth	797 Hz/Px

Sequence - Part 2

RF pulse type	Normal
Gradient mode	Fast
RF spoiling	On

Sequence - Assistant

Mode	Off
------	-----