

| | |
|--|--------------------------------|
| Policy Title: Breast Screening Standard Digital Mammography Digitized Image Display | |
| Section: Quality Management | Reference No. SG-DG 900 |
| Effective: February 2013 | Revision: February 2018 |

1. SCOPE

Breast Screening Centre Staff
Breast Screening Program Chief Radiologists
Breast Screening Program Chief Technologists
Breast Screening Quality Assurance Support Group
Radiology Managers

2. POLICY

All digitized images of prior screening films must be displayed on the review workstation with proper orientation, patient information, and correctly labelled view positioning (RCC, LCC, RMLO, and LMLO). The hanging protocols should display images right to left to correspond with the layout on the Breast Screening mammography report form.

Digitizer software used for digitizing prior screening mammography films shall support the IHE Mammography Image Profile as the “Acquisition Modality” Actor.

Digitizer used for digitizing prior screening mammography films must have been certified for mammography and must comply with the specification requirement and quality assurance program as recommended by the Breast Screening Quality Assurance Support Group.

3. RELATED POLICIES

[SG-DG 400 – Breast Screening Standard for Soft Copy Display \(Workstations\)](#)
[SG-DG 600 – Breast Screening Standard Digital Mammography Interpretation Image Display](#)

4. RESPONSIBLE PARTY

Professional Practice Leader – Breast Screening Technologists
Breast Screening Quality Assurance Support Group