

When the AI result arrives

Arrival of the AI result relative to the radiology report. Observational study, 20 centers, 21,872 exams with a linked report.

■ Before report creation ■ During reporting ■ After finalization ("Too Late" rate)
▨ During or after finalization (group rows: reported combined)

All exams

n = 21,872



Radiography

n = 15,260



CT and MRI

n = 5,246



80.6% of AI results were ready before the report was created, but 7.2% arrived after finalization. CT and MRI results arrived during or after finalization more often than radiography (29.6% vs 13.6%). Fix: push routing, integration and per-modality time-to-display targets, starting with CT and MRI.

Group rows: the paper reports the share during or after finalization combined. Mammography (n = 1,366) is in neither group.

Morozov S, Heracleous N, Korka D, Thoully C, Dufour B, Novarina O, Rizk B. AI Latency, Report Turnaround Time, and Adoption in a Multi-Vendor AI Ecosystem: A Multi-Site Observational Study. J Am Coll Radiol. 2026. doi:10.1016/j.jacr.2026.09.026

Redrawn from the authors' data

CC BY-NC-ND 4.0 · share unchanged with credit · no commercial use



AI in radiology: what works, what to fix

Observational study, 20 centers, 4.5 years, ~389,000 AI-processed exams

91%

of radiologists adopted AI

-26%

report time in trauma
radiography with integrated AI
(radiologist-adjusted)

80.6%

of AI results ready before
reporting; 7.2% arrived after
finalization

72%

of AI delay is data routing: fix
with push routing and
integration

Observational study: associations, not causation.

Morozov S, Heracleous N, Korka D, Thouly C, Dufour B, Novarina O, Rizk B. AI Latency, Report Turnaround Time, and Adoption in a Multi-Vendor AI Ecosystem: A Multi-Site Observational Study. J Am Coll Radiol. 2026. doi:10.1016/j.jacr.2026.09.026

Redrawn from the authors' data

CC BY-NC-ND 4.0 · share unchanged with credit · no commercial use



AI in radiology: what works, what to fix

Observational study, 20 centers, 4.5 years

91% of radiologists adopted AI

-26% report time, trauma
radiography

80.6% ready before reporting; 7.2%
after finalization

72% of delay is routing: fix with
push routing

Morozov S, Heracleous N, Korka D, Thouly C, Dufour B, Novarina O, Rizk B. AI Latency, Report Turnaround Time, and Adoption in a Multi-Vendor AI Ecosystem: A Multi-Site Observational Study. J Am Coll Radiol. 2026.
doi:10.1016/j.jacr.2026.09.026

Redrawn from the authors' data

CC BY-NC-ND 4.0 · share unchanged with credit · no commercial use



aimonitoring.drsergeymorozov.com

Monitoring that helps AI deliver: three layers

What was measured across 20 centers over 4.5 years, and the loop that acts on it. Observational study.

1 Machine time

- Total latency, image to AI result
- Data routing vs AI processing
- Quarterly latency stability per solution

In this study

Median 2.06 min; 72% was data routing; five of nine solutions had stable quarterly latency, four did not

2 Workflow time

- AI arrival vs report start and sign-off
- Share ready before reporting; Too Late rate
- Report time with vs without AI

In this study

80.6% ready before reporting, 7.2% after finalization; report time -26% trauma XR, -18% knee MRI (adjusted)

3 Radiologist perception

- Trust, quality, productivity (1 to 5)
- Net Promoter Score
- Survey radiologists on a fixed schedule (this study: two waves in 2025)

In this study

91.4% active adoption (53/58); perceived productivity 2.94 of 5, below neutral despite measured savings



Plan

Set time-to-display targets before go-live

Do

Push AI results automatically into the reading workflow

Check

Review latency, stability and radiologist scores every quarter

Act

Upgrade routing; replace tools that do not deliver

Morozov S, Heracleous N, Korka D, Thoully C, Dufour B, Novarina O, Rizk B. AI Latency, Report Turnaround Time, and Adoption in a Multi-Vendor AI Ecosystem: A Multi-Site Observational Study. J Am Coll Radiol. 2026. doi:10.1016/j.jacr.2026.09.026

Redrawn from the authors' data

CC BY-NC-ND 4.0 · share unchanged with credit · no commercial use

