

FDA 510(k) Submission

Device: GlucoMonitor Pro 2000

Date: 2023-09-10

Author: Regulatory Affairs Team

Device Description:

- Name: GlucoMonitor Pro 2000
- Components: Includes CMP-2002 (Adhesive Strip)
- Classification: Class II medical device
- Indications for Use: Continuous glucose monitoring for diabetic patients

Regulatory Notes:

- Component-Specific Notes: Adhesive (CMP-2002) tested for biocompatibility and durability under FDA protocols
- Conditions for Approval: Any changes to CMP-2002 must be reviewed under a new 510(k) submission or pre-approval amendment