

IT'S THE TIME TO DISCO

- Leverages existing TOPUP capability to train
- Much faster approach



HHS Public Access

Author manuscript

Magn Reson Imaging. Author manuscript; available in PMC 2020 December 01.

Published in final edited form as:

Magn Reson Imaging. 2019 December ; 64: 62–70. doi:10.1016/j.mri.2019.05.008.

Synthesized b0 for diffusion distortion correction (Synb0-DisCo)

Kurt G Schilling¹, Justin Blaber³, Yuankai Huo², Allen Newton^{1,5}, Colin Hansen², Vishwesh Nath³, Andrea T. Shafer⁴, Owen Williams⁴, Susan M. Resnick⁴, Baxter Rogers¹, Adam W Anderson¹, Bennett A Landman^{1,2,3}

¹Vanderbilt University Institute of Imaging Science, Vanderbilt University, Nashville, TN

²Department of Electrical Engineering, Vanderbilt University, Nashville, TN

³Electrical Engineering & Computer Science, Vanderbilt University, Nashville, TN

⁴Laboratory of Behavioral Neuroscience, National Institute on Aging, National Institutes of Health, Baltimore, MD

⁵Department of Radiology and Radiological Sciences, Vanderbilt University Medical Center, Nashville, TN

Abstract

Diffusion magnetic resonance images typically suffer from spatial distortions due to susceptibility induced off-resonance fields, which may affect the geometric fidelity of the reconstructed volume and cause mismatches with anatomical images. State-of-the art susceptibility correction (for example, FSL's TOPUP algorithm) typically requires data acquired twice with reverse phase

Synb0-DisCo

