

Image Acquisition Parameters — Screen Instructions & Validation Rules

SYNTHETIC TEST DATA — This document was generated for RAG system development and testing only. It does not represent real clinical guidance, a real hospital policy, or real patient information. Do not use for actual patient care.

Document ID: DOC_028 | Type: screen_instructions | Domain: radiology | Workflow: WF_001 | Screen: SCREEN_04

Purpose

This screen captures the information needed to progress the "Image Acquisition Parameters" step of workflow WF_001.

Fields on This Screen

Field	Required
Protocol Name	Yes
Slice Thickness	Yes
Field of View	Yes
Technologist ID	Yes

Required Verifications

#	Required Verification
1	Confirm the acquisition protocol matches the clinical indication
2	Confirm technologist ID is captured for the acquisition record
3	Confirm image count matches the expected series count for the protocol

Common Errors & Recovery

Common Error	Likely Cause
Protocol mismatch versus clinical indication	Data entry or workflow-timing issue
Incomplete series (image count below expected minimum)	Data entry or workflow-timing issue

Screen Reference Mockup (Synthetic)

SCREEN_04 — Image Acquisition Parameters (SYNTHETIC MOCKUP)

Workflow WF_001 > Image Acquisition Parameters

Protocol Name:	
Slice Thickness:	
Field of View:	
Technologist ID:	SYN-7468

Required Verifications

- ☐ Confirm the acquisition protocol matches the clinical indi...
- ☐ Confirm technologist ID is captured for the acquisition re...
- ☐ Confirm image count matches the expected series count for ...

Save	Back
------	------

synthetic mockup - not a real application screen