

<div>Calibration Certificate of PCD Gauge</div>			<div>Certificate No. : SAF-CAL2024-38 Date of Issue : 26-02-2024 Date of Calibration : 26-02-2024 Next Calibration Due : 03-03-2024</div>				
<div>Name and Address of Customer: M/s. Super Auto Forge Private Limited Vallam - Oragadam</div>			<div>Description of Gauge / Instrument : Range/Size : 25.6 mm Identification No. : S-Test SI No. : 1256 Make : Metric</div>				
Environmental Condition							
Temperature		20±2°C		Humidity		30% to 75%	
Standard Ref / SOP No.				IS 3455 / SOP-2			
Master Used for Calibration							
SiNo	IMTE No.	Item Description	Certificate No.	Valid Upto	Traceability		
1	UM-700-YE-112	2D Height Gauge	undefined	08-01-2025	Electronica Mechatronic Systems (I) Pvt. Ltd.		
Calibration results							
Parameter	Nominal Size	Permissible Size (mm)		Observed Size (mm)	Cal Status		
		Min/Max	Wear Limit				
Dia 1	23.2	23.21/23.229		23.215	ok		
Dia 2	23.5	23.5/23.52		23.515	ok		

Remarks :

- Acceptance Remarks : accepted
- The Expanded uncertainty is ± 0.002 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★