

<b>Calibration Certificate of Comparator Stand</b>			<b>Certificate No.</b> : SAF-CAL2024-25 <b>Date of Issue</b> : 29-02-2024 <b>Date of Calibration</b> : 29-02-2024 <b>Next Calibration Due</b> : 14-03-2024		
<b>Name and Address of Customer:</b> M/s. Super Auto Forge Private Limited Vallam - Oragadam			<b>Description of Gauge / Instrument :</b> <b>Range/Size</b> : 200X200 mm <b>Identification No.</b> : TEST-1 <b>SI No.</b> : 147 <b>Make</b> : Alfa		
<b>Environmental Condition</b>					
<b>Temperature</b>		20±2°C		<b>Humidity</b> : 30% to 75%	
<b>Standard Ref / SOP No.</b>			IS 3456 / Saf-06		
<b>Master Used for Calibration</b>					
<b>SiNo</b>	<b>IMTE No.</b>	<b>Item Description</b>	<b>Certificate No.</b>	<b>Valid Upto</b>	<b>Traceability</b>
1	I-CMM-1	CMM	undefined	16-10-2024	undefined
<b>Calibration results</b>					
<b>Parameter</b>	<b>Nominal Size (mm)</b>	<b>Permissible Size (mm)</b>		<b>Observed Size (mm)</b>	<b>Cal Status</b>
		<b>Min/Max</b>			
Flatness	200X200	0/0.005		0.003	ok

Remarks :

- Acceptance Remarks : accepted
- The Expanded uncertainty is ± 0.006 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★