



<b>Calibration Certificate of Thread Depth Gauge</b>			<b>Certificate No.</b> : CAL2024-55 <b>Date of Issue</b> : 09-03-2024 <b>Date of Calibration</b> : 09-03-2024 <b>Next Calibration Due</b> : 23-03-2024		
<b>Name and Address of Customer:</b> M/s. ABI Soorai Soorai - Soorai			<b>Description of Gauge / Instrument :</b> <b>Range/Size</b> : 13.60 MIN M14X1.50P-6G mm <b>Identification No.</b> : 11GS5190-A20 <b>SI No.</b> : - <b>Make</b> : -		
<b>Environmental Condition</b>					
<b>Temperature</b>		20±2°C		<b>Humidity</b>	
				30% to 75%	
<b>Standard Ref / SOP No.</b>			IS 3455 / SOP-1		
<b>Master Used for Calibration</b>					
<b>SiNo</b>	<b>IMTE No.</b>	<b>Item Description</b>	<b>Certificate No.</b>	<b>Valid Upto</b>	<b>Traceability</b>
1	A-685	2D Microhite	A080/2018	09-07-2024	undefined
<b>Calibration results</b>					
<b>Parameter</b>	<b>Nominal Size (mm)</b>	<b>Permissible Size (mm)</b>		<b>Observed Size (mm)</b>	<b>Cal Status</b>
		<b>Min/Max</b>	<b>Wear Limit</b>		
Go Dia	13.60	13.61/13.63	13.60	13.615	ok

Remarks :

- Acceptance Remarks : accepted
- The Expanded uncertainty is ± 0.002 mm at 95.45% confidence level with coverage factor k=2.
- This certificate is digitally signed and does not require any seal or signature

★★★ END OF THE CERTIFICATE ★★★

Dharani  
Calibrated By

Dharani  
Approved By